



science and policy
for a healthy future

HORIZON2020 Programme
Contract No. 733032 HBM4EU

ICI / EQUAS REPORT

DINCH/round_02 (2018)

DINCH biomarkers in urine

Version / date of issue	1 / 10-04-2019
Organiser	RIKILT – Wageningen University & Research Akkermaalsbos 2 6708 WB, Wageningen The Netherlands
Coordinator	Ingrid Elbers, ingrid.elbers@wur.nl
Author(s) (Short name of institute)	Ingrid Elbers (RIKILT), Hans Mol (RIKILT)
Approved by:	Hans Mol (RIKILT)

Table of contents

Table of contents	2
1 Summary	3
2 Introduction	4
2.1 Confidentiality	4
3 Control material	5
3.1 Preparation of control material	5
3.2 Homogeneity of control material	5
3.3 Stability of control material	5
4 Organisational details	6
4.1 Participants	6
4.2 Dispatch and instructions	6
4.3 Deviations from ICI/EQUAS SOPs	6
5 Data evaluation	7
5.1 False positives and <LOQ	7
5.2 Assigned value	7
5.3 Target standard deviation (σ_T)	7
5.4 ICI/EQUAS standard deviation (RSD_R)	7
5.5 Z-scores	8
6 Results and discussion	9
6.1 Results submitted by participants	9
6.2 Assigned values and (target) standard deviations	9
6.3 Assessment of laboratory performance	10
6.4 Conclusions and recommendations	10
7 References	11

Appendices

- 1 List of participants
- 2 Homogeneity data
- 3 Copy of letter of invitation
- 4 Copy of letter/instructions sent together with test samples
- 5 Scope and LOQs of the laboratories
- 6 Assigned values and Z-scores
- 7 Graphical representation of Z-scores

1 Summary

Within the frame of the HBM4EU project, an EQUAS study was organised on the determination of two DINCH biomarkers in urine. This was the second ICI/EQUAS round for this substance group within the HBM4EU program.

In total 14 laboratories were invited for this second round, of which 12 laboratories registered. Results were received from all 12 laboratories, located in 8 EU countries and the USA (see Appendix 1).

In December 2018, each participant received two burdened control materials of human urine, A and B (single tube each), containing DINCH biomarkers in the range 4-23 ng/ml.

Homogeneity assessment showed that both materials were sufficiently homogeneous for ICI/EQUAS testing. The stability test demonstrated no significant loss of the biomarkers during the course of the EQUAS test.

The proficiency of the laboratories was assessed through Z-scores calculated using the mean concentration as established by expert laboratories as assigned value, and a fixed fit-for-purpose relative target standard deviation (FFP-RSD_R) of 25%. Assigned values and Z-scores could be determined for both biomarkers in both test materials.

One laboratory determined only one of the two biomarkers, all others determined both. The percentage of satisfactory Z-scores was around 80% for both biomarkers when including the results from the expert laboratories.

The characteristics and outcome of this EQUAS are summarized in Table 1.

Table 1. Summary table EQUAS results.

Biomarker	Sample	Assigned value (ng/ml)	study RSD _R ¹⁾ %	Z-scores			
				No	satisfactory	questionable	unsatisfactory
OH-MINCH	A	6.91	51%	12	75%	17%	8%
	B	23.0	49%		83%	8%	8%
cx-MINCH	A	3.66	55%	11	82%	9%	9%
	B	12.1	70%		82%	9%	9%

¹⁾ interlaboratory relative standard deviation (robust RSD based on participants' results, excluding expert labs)

2 Introduction

Interlaboratory Comparison Investigations (ICI) and External Quality Assurance Schemes (EQUAS) are tools to access the proficiency of laboratories, and the comparability and reliability of analytical methods. Participation in ICI / EQUAS forms an integral part of quality control, in addition to initial and on-going in-house method validation.

This EQUAS study has been organised within the frame of HBM4EU as part of the Quality Assurance program for biomonitoring analyses, following protocols HBM4EU-SOP-QA-001 to 004 which are available through the HBM4EU website. Within HBM4EU, participation in ICI/EQUAS exercises is mandatory for laboratories that will analyse HBM4EU samples.

This report describes the outcome of the 2nd ICI/EQUAS round for DINCH in urine and was organised by RIKILT – Wageningen University & Research in The Netherlands. RIKILT is ISO/IEC 17043 accredited for organisation of proficiency tests, but the specific substances in this EQUAS study were outside the specified scope of accreditation.

The selection of the most relevant/feasible biomarkers for DINCH was previously done in WP9, and has been described in Deliverable report 9.2 v1.1. Based on this, two target biomarkers were included in the EQUAS for DINCH biomarker analysis (see Table 2).

Table 2. Biomarkers for DINCH* included in the EQUAS.

Biomarker	
OH-MINCH	cyclohexane-1,2-dicarboxylate-mono-(7-hydroxy-4-methyl)octyl ester
cx-MINCH	cyclohexane-1,2-dicarboxylate-mono-(7-carboxylate-4-methyl)heptyl ester

* Di-isononyl cyclohexane-1,2-dicarboxylate

For this second round, the concentrations aimed at were between the median and roughly 1.5 times the 95th percentile concentrations as reported by [Correia-Sá 2017]. These relatively high concentrations (4-23 ng/ml) were chosen given the particular difficulties of this analysis, as observed in the first round.

The LOQs provided by the participants during registration for the EQUAS ranged from 0.1 to 1 µg/L).

2.1 Confidentiality

In this report the identity of the participants and the information provided by them are treated as confidential. However, lab codes of the participants will be disclosed to the HBM-QAU for performance assessments.

3 Control material

3.1 Preparation of control material

For this EQUAS two control materials, A and B, were prepared, one aiming at concentrations in the range of approximately 5 ng/ml and one roughly three to four times higher. The control materials were prepared by blending aliquots of different burdened human urine samples. The burdened human urines were kindly provided by the Institute for Prevention and Occupational Medicine of the German Social Accident Insurance (IPA), with concentration estimates.

For blending, the selected materials were thawed, the appropriate volumes taken and mixed. The blend (approx. 500 ml) was centrifuged to remove any precipitates. Then the urine was aliquoted (4 ml portions) into coded polypropylene tubes with screwcap. The tubes were stored in the freezer (28 November 2018, <-18°C). Part of the tubes were stored at -80°C as reference for stability testing.

3.2 Homogeneity of control material

Homogeneity testing was done as described in HBM4EU-SOP-QA-002. Ten tubes of control material A and ten tubes of control material B were randomly selected from the freezer and sent to IPA for analysis. Each sample was analysed in duplicate. In brief, after thawing/mixing, an aliquot of the urine was taken, isotope labels of the biomarkers were added as internal standard, and a deconjugation step using *E. coli* β -glucuronidase was performed. The deconjugated urine was analysed by on-line SPE coupled to LC-MS/MS. The analysis results were sent to the organiser and processed according to the SOP using an Excel macro ("HBM4EU macro homogeneity test v1.xlsm"). The mean concentrations and relative standard deviations as obtained during homogeneity testing are presented in Table 3. The statistical evaluation of materials A and B for each of the biomarkers are provided in Appendix 2. It was concluded that homogeneity was adequate for both biomarkers in both control materials.

Table 3. Concentration of DINCH biomarkers as obtained during homogeneity testing (details see Appendix 2).

Biomarker	material A		material B	
	$\mu\text{g/L}$	RSD _r	$\mu\text{g/L}$	RSD _r
OH-DINCH	6.91	2%	22.96	1%
cx-DINCH	3.53	2%	12.05	2%

3.3 Stability of control material

Stability testing was done according to HBM4EU-SOP-QA-002. At the day of preparation of the control materials, randomly selected test samples of material A and B were stored at -80°C. The assumption here is that under these conditions, the biomarkers are stable in urine. After the (extended) deadline of submission of analysis results by the participants, six test samples of both material A and B stored at -80°C, and six samples of material A and B randomly selected from the -18°C freezer where the EQUAS samples were stored, were sent to IPA for analysis (method same as described in 3.2). Results were sent to the organiser to assess stability. The control materials were considered stable when the difference of the means of the -80°C and the -18°C samples was $\leq 0.3 \sigma_T$. In both control materials this was the case for both biomarkers.

EQUAS report DINCH round 2	Version: 1	Date: 10-04-2019	Page: 6
DINCH biomarkers in urine, Round-2			

4 Organisational details

4.1 Participants

Participants for this EQUAS study were laboratories from the HBM4EU consortium (including linked-third parties) that had been included as candidate laboratories for analyses in the frame of the HBM4EU project through WP9 (Task 9.2, Deliverable 9.3). A list of 14 eligible candidate laboratories was provided to RIKILT. Invitation letters were sent by e-mail on 26 November 2018 (see Appendix 3). For registration, each participant was asked to provide whether or not both target biomarkers were included in their scope of analysis, and the LOQs in µg/L (=ng/ml).

In total 12 laboratories from 8 EU countries and the USA registered. This included three expert laboratories, two from the HBM4EU consortium and one from the USA. Results were received from all laboratories (see Appendix 1).

4.2 Dispatch and instructions

Test materials (one tube A and one tube B, with unique codes, containing approx. 4 ml urine each, frozen conditions) were dispatched to the participants on 17th December 2018. The samples were packed in an insulation box with dry ice and sent by courier. Instructions and an "acknowledgement of receipt form" were included in the box and also sent by e-mail at the day of shipment (see Appendix 4). Participants were asked to check the content of the box upon receipt, to store the samples in the freezer, and to analysis the samples according to their routine method. The deadline for submission of results was initially 15th January 2019, but extended to 29th of January 2019 because the samples were shipped later than intended and because of the Christmas holidays.

In this round, special instructions were given to all laboratories regarding the transition to be used for quantification in the LC-MS/MS analysis, and the use of sufficiently wide acquisition windows to ensure that all isomer peaks of the biomarkers were included in the measurement (see also Appendices 2 and 3). This was done because DINCH biomarkers in burdened urine are isomeric mixtures which may result in multiple and/or broad peaks, and because the transition used for quantification may affect the analysis result. The very high variability of the results observed in the 1st round was attributed to this, and it was decided to harmonise on the quantifier transitions in order to improve interlaboratory precision.

Together with the instructions sent by email, also a request to provide detailed method information in an Excel file was sent to the participants. In this sheet, the participants were asked to specify the limit of quantification (LOQ) for each of the biomarkers. In addition, details on enzymatic deconjugation, cleanup, analysis technique, internal standards used, and precision data was asked for.

4.3 Deviations from ICI/EQUAS SOPs

For this 2nd round, the HBM4EU-QA-SOPs (version 2) were followed. There were no deviations from these SOPs.

5 Data evaluation

5.1 False positives and <LOQ

Classification of false positives and biomarkers reported as "<LOQ-value" or "not detected" is described in HBM4EU-SOP-QA-003. In this EQUAS there were no false positives and no non-detects. Therefore no further description is given here.

5.2 Assigned value

For EQUAS studies, the concentration as established by expert laboratories was used as assigned value. Expert laboratories were selected by the HBM4EU quality assurance unit. In this EQUAS round, a list of three laboratories was provided to the organiser, two from the HBM4EU consortium and one from the USA. The expert laboratories all agreed to collaborate. The expert laboratories received the same control material and instructions as the participants. However, instead of one test sample and single analysis, expert laboratories received six test samples to be analysed in duplicate. Upon receipt of their results and method information, the acceptability of the results for establishment of the expert value was verified. The following aspects were taken into account:

- precision (RSD_r) of the results provided by each expert lab.
- use of the isotopically labelled analogue as internal standard for each of the biomarkers analysed. For determination of the expert value, not using such internal standard was an exclusion criterion.
- use of the prescribed quantifier transitions for quantification (m/z 313>153 for OH-MINCH, m/z 327>173 for cx-MINCH). Not using the prescribed transitions was an exclusion criterion.

Next, the expert value was determined as described in HBM4EU-QA-001. In brief, using the individual means of the expert laboratories, the mean of the means was calculated and its relative uncertainty. The mean of means was used as assigned value when the relative uncertainty was below $0.7 \cdot \sigma_T$. When this condition was not met, and no outliers could be identified, then the uncertainty of the expert-derived mean was considered too high to be used as assigned value. The other requirement to be met was that the number of (remaining) individual expert means had to be at least three.

In case no expert value could be obtained, the consensus value derived from the combined results from both participants and expert laboratories could be used as an alternative.

5.3 Target standard deviation (σ_T)

For calculation of the Z-scores, a fit-for-purpose relative target standard deviation (FFP- RSD_R) of 25% of the assigned value was used as target standard deviation for proficiency. This was the default indicated in HBM4EU-SOP-QA-003 and considered appropriate based on the outcome of the 1st round.

5.4 ICI/EQUAS standard deviation (RSD_R)

To gain insight in the actual interlaboratory variability of the biomarkers determined in this study, the robust relative standard deviation (RSD_R) was calculated based on the participants' results, as described in HBM4EU-SOP-QA-003. For this the results of the expert laboratories were not included.

5.5 Z-scores

Z-scores were calculated according to SOP HBM4EU-SOP-QA-003.

$$Z = \frac{x - C}{\sigma_T} \quad (1)$$

with: Z = Z-score for the submitted analysis result;
x = result submitted by the laboratory;
C = expert-assigned value;
 σ_T = target standard deviation, here 0.25*C

In accordance with ISO 13528 and ISO 17043 and the deliverable D 9.4 “*The Quality Assurance/Quality Control Scheme in the HBM4EU project*”, Z-scores are classified as presented in Table 4.

Table 4: Classification of Z-scores

$ Z \leq 2$	Satisfactory
$2 < Z < 3$	Questionable
$ Z \geq 3$	Unsatisfactory

EQUAS report DINCH round 2	Version: 1	Date: 10-04-2019	Page: 9
DINCH biomarkers in urine, Round-2			

6 Results and discussion

6.1 Results submitted by participants

In total 12 laboratories from eight EU countries and the USA (see appendix 1) agreed to participate in this EQUAS. All submitted results.

The scope of the laboratories and the LOQs for both biomarker as submitted together with the analysis results through the Excel method information sheet are provided in Appendix 5. Both biomarkers were measured by all laboratories with one exception.

The LOQs were generally in the range 0.2-0.5 ng/ml (see Appendix 5).

The individual analysis results of the laboratories are included in appendix 6.

Laboratories were asked to provide details on the method used for analysis. In general, the laboratories did not do any filtration/centrifugation after thawing the urine sample, added isotope label(s) to an aliquot of 0.2-3 ml urine, and adjusted the pH to values ranging from 4.5 to 6.5. All labs did an enzymatic deconjugation step, mostly using *E. Coli* β -glucuronidase (one lab used *Helix Pomatia* β -glucuronidase/aryl-sulfatase), at 37°C for 1.5-2.5 hours (in some cases overnight). In most cases the deconjugated urine was acidified and then extracted/preconcentrated using on-line or off-line SPE. The biomarkers were then measured using liquid chromatography with mass spectrometric detection (LC-MS/MS, electrospray ionisation in negative mode). In most cases, the laboratories used the isotopically labelled analogue of the biomarker as internal standard. With only one exception, the prescribed quantifier transitions were used for quantification.

In the method information sheet, the laboratories were also asked to provide existing precision data from (on-going) validation, i.e. repeatability, intermediate precision and measurement uncertainty. While most laboratories provided repeatability data, less information on intermediate precision was provided. Only few laboratories provided data on measurement uncertainty.

6.2 Assigned values and (target) standard deviations

The assigned value was the expert-assigned value as derived from replicate analysis of the control materials by three expert laboratories as described in 5.2. The repeatability of the results for each biomarker by each expert lab was very good (typically $RSD_r < 6\%$). In all cases, the isotopically labelled analogue of the biomarker was used as internal standard, and the prescribed transition for quantification. The individual means of the expert labs were in very good agreement with each other.

Expert-assigned values could be established for both biomarkers in both control materials. The assigned values and their uncertainties are included Appendix 6.

The target standard deviation used for determination of the Z-scores was 25% (0.25**C*) (see 5.3 and 5.5). To verify how this fixed target value compares to the actual interlaboratory variability of the results, the relative standard deviation (study RSD_R , robust statistics) derived from the participants' results (excluding the results from the expert labs) were calculated. The RSD_R 's are included in Appendix 6. Despite the harmonisation of the quantifier transitions and recommendations provided in a webinar after the first round, they were still high, around 50% and 70% for cx-MINCH in material B. Because the number of laboratories was rather small (12), the exclusion of the results of the expert laboratories had a high impact on the RSD_R of the remaining nine participants.

EQUAS report DINCH round 2	Version: 1	Date: 10-04-2019	Page: 10
DINCH biomarkers in urine, Round-2			

6.3 Assessment of laboratory performance

Z-scores were calculated for the two biomarkers in both control materials. For each of the laboratories, the individual Z-scores for the biomarkers in both samples are provided in Appendix 6. A graphical representation of the individual Z-scores is shown in appendix 7.

A summary of number of laboratories that reported results, and the percentage of satisfactory, questionable, and unsatisfactory Z-scores is included in Table 1. The percentage of satisfactory Z-scores obtained was 75%-83%.

6.4 Conclusions and recommendations

In this 2nd ICI/EQUAS round on two DINCH biomarkers in urine, 12 laboratories (including three expert labs) registered and submitted results.

The interlaboratory variability of results was high, despite standardization of quantifier transitions and other recommendations done during a webinar after the first round. However, due to the use of expert-assigned values, it was possible to assess the proficiency of the laboratories through z-scores. Apart from the expert laboratories, five out of the nine other participants achieved satisfactory results for both biomarkers in both control materials. In this sense, progress has been made compared to the first round.

The participants with questionable or unsatisfactory results are recommended to do a root cause analysis to find the reason for the deviating results, and seek assistance from HBM4EU expert laboratories if needed.

EQUAS report DINCH round 2	Version: 1	Date: 10-04-2019	Page: 11
DINCH biomarkers in urine, Round-2			

7 References

Analytical Methods Committee, 1989a, Robust statistics - How not to reject outliers Part 1. Basic concepts, Analyst, 114, 1693-1697.

Analytical Methods Committee, 1989b, Robust statistics - How not to reject outliers Part 2. Interlaboratory trials, Analyst, 114, 1699-1702.

Correia-Sá L, Schutze A, Norberto S, Calhau C, Domingues VF, Koch HM, Environment Int. 102 (2017) 79-86.

HBM4EU-SOP-QA-001 (version 2) "Organisation of Interlaboratory Comparison Investigations (ICI) and External Quality Assurance Schemes (EQUAS) of interlaboratory studies"

HBM4EU-SOP-QA-002 (version 2) "Preparation of control materials for Interlaboratory Comparison Investigations (ICI) and External Quality Assurance Schemes (EQUAS)"

HBM4EU-SOP-QA-003 (version 2) "Evaluation of results from Interlaboratory Comparison Investigations (ICI) and External Quality Assurance Schemes (EQUAS)"

HBM4EU-SOP-QA-004 (version 2) "Reporting of results of Interlaboratory Comparison Investigations (ICI) and External Quality Assurance Schemes (EQUAS)"

Note: the above mentioned SOPs can be found on the HBM4EU website:

<https://www.hbm4eu.eu/online-library/?mdocs-cat=mdocs-cat-1&mdocs-att=null>

HBM4EU Deliverable 9.2 Prioritised list of biomarkers, matrices and analytical methods for the 1st prioritisation round of substances. <https://www.hbm4eu.eu/deliverables/>

ISO/IEC 17043:2010, Conformity assessment – General requirements for proficiency testing

ISO 13528, 2015, Statistical methods for use in proficiency testing by interlaboratory comparison.

Official Methods of Analysis Program Manual, 2002, Appendix D: Guidelines for Collaborative Study Procedures To Validate Characteristics of a Method of Analysis. Association Of Analytical Communities International. http://www.aoac.org/vmeth/Manual_Part_6.pdf.

Appendix 1. List of countries participating


Country	Number of laboratories participating
Belgium	2
Czech Republic	2
Denmark	1
Germany	1
Greece	1
Hungary	1
Norway	1
Slovakia	1
Sweden	1
USA	1

Appendix 2. Homogeneity data


	Control material A		Control material A	
	OH-MINCH		cx-MINCH	
	replicate-1	replicate-2	replicate-1	replicate-2
1	6.64	6.83	3.51	3.57
2	6.91	7.04	3.53	3.62
3	6.83	6.90	3.51	3.63
4	6.71	6.76	3.49	3.38
5	6.79	6.92	3.51	3.51
6	6.91	7.12	3.54	3.57
7	6.94	7.03	3.43	3.56
8	7.03	7.06	3.47	3.63
9	6.95	6.93	3.52	3.46
10	6.87	7.09	3.58	3.53
grand mean	6.913		3.528	
Stdev	0.128		0.064	
VC%	2%		2%	
Cochran's test				
C	0.270		0.292	
Ccrit	0.602		0.602	
C < Ccrit →	No outliers detected		No outliers detected	
target σ_{FFP}	1.728		0.882	
$s_x =$	0.1111		0.0442	
$s_w =$	0.0947		0.0662	
$s_s =$	0.0887		0.0000	
critical= $0.3\sigma_{FFP}$	0.5185		0.2646	
$s_s < \text{critical?}$	Homogeneity adequate		Homogeneity adequate	
$s_w < 0.5*\sigma_{FFP}?$	Method suited		Method suited	

	Control material B		Control material B	
	OH-MINCH		cx-MINCH	
	replicate-1	replicate-2	replicate-1	replicate-2
1	23.34	23.30	12.31	12.46
2	23.25	23.12	11.92	12.43
3	23.05	22.81	12.28	12.20
4	23.00	22.56	12.31	12.10
5	23.00	23.04	12.21	11.95
6	22.26	22.55	11.81	12.06
7	23.47	23.10	11.81	12.16
8	22.31	23.20	11.97	11.65
9	22.70	23.03	11.87	11.72
10	23.17	22.89	11.98	11.76
grand mean	22.958		12.048	
Stdev	0.333		0.241	
VC%	1%		2%	
Cochran's test				
C	0.538		0.343	
Ccrit	0.602		0.602	
C < Ccrit →	No outliers detected		No outliers detected	
target σ_{FFP}	5.739		3.012	
$s_x =$	0.2762		0.2004	
$s_w =$	0.2713		0.1948	
$s_s =$	0.1988		0.1455	
critical= $0.3\sigma_{FFP}$	1.7218		0.9036	
$s_s < \text{critical?}$	Homogeneity adequate		Homogeneity adequate	
$s_w < 0.5*\sigma_{FFP}?$	Method suited		Method suited	

Appendix 3. Copy of letter of invitation



WAGENINGEN
UNIVERSITY & RESEARCH



science and policy
for a healthy future

Postbus 320 • 6700 AH WAGENINGEN The Netherlands

HBM4EU: Announcement / invitation to participate in ICI-EQUAS study DINCH/round 02

Title of ICI: DINCH biomarkers in urine

Dear colleague,

Within the frame of HBM4EU, RIKILT – Wageningen University & Research announces the second round of ICI/EQUAS for the determination of DINCH-biomarkers in urine. The aim of ICI/EQUAS exercises is to provide laboratories with an assessment of their analytical performance and reliability of their data in comparison with other laboratories and/or expert laboratories. This will aid in the quality improvement of analysis in human biomonitoring at each of the laboratories.

Participation is mandatory for laboratories analysing samples in the frame of HBM4EU.

1. Samples

- Two urine samples (approximately 2-4 ml) will be supplied for the analysis of DINCH biomarkers. See the registration form for potentially present biomarkers.
- The samples are expected to be sent by mid December 2018. The distribution of the samples will be announced by e-mail.

2. Quantitative analysis


- The results have to be reported before Jan 15th 2019.
- Results must be submitted via the electronic submission form (<https://crlwebshop.wur.nl/apex/f?p=307:LOGIN>) for which, if not already done so for the first round, each participant must register, as explained in the document "Instruction registration login account RIKILT Laboratory Quality Services".

3. Report

- A report of the proficiency test will be dispatched in February 2019.
- Results of the proficiency test will be presented anonymously.

4. Additional information

- The analysis results submitted by the laboratory and the performance derived from that will be treated confidentially, and will never be disclosed to third parties (government, accreditation bodies) except the HBM4EU QAU, without permission of the laboratory.



RIKILT

DATE
November 28, 2018

SUBJECT
Invitation for ICI-EQUAS study
DINCH biomarkers in urine

FOCUS/ADDRESS
Afdelingslabo 2
6708 WB, Wageningen
The Netherlands

INTERNET
www.wur.nl/rikilt

DOC NUMBER
09095104

HANDLED BY
Ingrid Elbers

TELEPHONE
+31 317 481491

EMAIL
gi.rikilt@wur.nl

Wageningen Research
Foundation/RIKILT is part of
Wageningen University & Research.
RIKILT carries out research into the
safety and reliability of food and
feed. RIKILT is ISO 17025 and ISO
17043 accredited (the accredited
bodies are described on www.ris.nl)
(no. L014 and R012).

Appendix 3. Copy of letter of invitation (continued)

DATE
November 28, 2018

PAGE
2 of 2

- Please take advantage of the feedback and suggestions for performance improvement provided to the participants during the ~~Webex~~ on 11th October 2018. Slides on this have been sent to the participants, if you do not have them please contact us.
- DINCH biomarkers are isomeric mixtures typically resulting in multiple and/or broad peaks in real samples, and the transition used for quantification affects the analysis result. For this reason the participants are asked to use the following transitions for quantification:
OH-MINCH: m/z 313 > 153
 α -MINCH: m/z 327 > 173
- Also, please ensure the acquisition windows for these compounds are sufficiently wide (at least from 2 min before until 2 min after the retention time of the analytical standard, depending on your specific LC method even longer) to ensure all peaks belonging to the biomarker are measured.

5. Costs

- For partners and linked-third parties of HBM4EU, participation is free of charge. Please note that the participant is responsible for custom clearance and associated costs if applicable.

6. ~~Calendar~~

- Deadline ~~registration~~: 7th December 2018
- Distribution of samples: mid December 2018
- Deadline submission of results: 15th January 2019~~tion~~

If you would like to participate, please fill out the registration form and send it to me before 7th December 2018 by e-mail (pt.rikilt@wur.nl). Please indicate on the registration form which biomarkers are within the scope of your method, and the estimated LOQ.

Hoping to welcome you for this ICI/EQUAS round,

Yours sincerely,



Ingrid Elbers (organiser proficiency test/ICI/EQUAS)
Hans Mol (scientific expert)

Appendix 3. Copy of letter of invitation (continued)



Registration form ICI study DINCH/round_02 (RIKILT code 2018-17)

Contact person:
E-mail:
Username for RIKILT web application: (needed for reporting results)
No <u>username</u> ? Please register at: Register (Ctrl + click)

Please indicate below the biomarkers that are within the scope of your method, and provide an estimate of your LOQ. We would appreciate your registration for both biomarkers.

Biomarkers included in the scope of your method (please tick)	Abbreviation	LOQ (µg/l urine)
<input type="checkbox"/> cyclohexane-1,2-dicarboxylic mono hydroxyisooctyl ester	OH-MINCH	
<input type="checkbox"/> cyclohexane-1,2-dicarboxylic mono carboxyisooctyl ester	cx-MINCH	

I hereby accept the conditions for participation as outlined in the letter accompanying this form.


Date / Signature:

Please sign a print of this document and e-mail a scan to pt.rikilt@wur.nl.


Please subscribe before 7th December, 2018

Ingrid Elbers (organiser proficiency test/ICI/EQUAS)
Hans Mol (scientific expert)
RIKILT - Wageningen University & Research
The Netherlands

Appendix 4. Copy of letter/instructions sent together with test samples



Postbus 330 • 8700 AH Wageningen The Netherlands



science and policy
for a healthy future

RIKILT

DATE
December 17, 2018

SUBJECT
Instruction letter for ICI-EQUAS study DINCH biomarkers in urine

OUR REFERENCE
RIKILT/PT 2018-16

POSTAL ADDRESS
Aldkermadebos 2
8708 WB, Wageningen
The Netherlands

WEBSITE
www.wur.nl/rikilt

CAC NUMBER
09095104

HANDLED BY
Ingrid Elbers

TELEPHONE
+31 317 481451

EMAIL
pt.rikilt@wur.nl

Dear participant,


Thank you for participation in **HBM4EU ICI/EQUAS study DINCH/round_02** (RIKILT code 2018-17) for the determination of DINCH biomarkers in urine.

You will receive a parcel containing two randomly coded samples. Each sample contains approximately 4 ml of urine.

Please fill out the accompanied 'acknowledgement of receipt form' and return it immediately upon receipt of the samples, preferably by e-mail (pt.rikilt@wur.nl).

Instructions:

- Upon receipt, store the samples in the freezer until analysis.
- Before analysis, thaw and re-homogenize the samples according to your laboratory's procedure.
- Please carry out a **single analysis** for each sample using the same procedure as used for analysis of samples in the frame of HBM4EU.
- Report the results in µg/L.
- DINCH biomarkers are isomeric mixtures typically resulting in multiple and/or broad peaks in real samples, and the transition used for quantification affects the analysis result. For this reason the participants are asked to use the following transitions for quantification:
OH-MINCH: m/z 313 > 153 ox-MINCH: m/z 327 > 173
- Also, please ensure the acquisition windows for these compounds are sufficiently wide (at least from 2 min before until 2 min after the retention time of the analytical standard, depending on your specific LC method even wider) to ensure all peaks belonging to the biomarker are measured.
- The deadline for submitting the results for this ICI/EQUAS is **Jan 29th, 2019**
- Please use the web application for entering your results (<https://crlwebshop.wur.nl/apex/f?p=307:LOGIN>). Information about the use of this web application was sent to you earlier by email.
- Your username is:
- Your password is:
- Your lab code to enter this proficiency test is:



100 years
1918 — 2018

Wageningen Research Foundation/RIKILT is part of Wageningen University & Research. RIKILT carries out research into the safety and reliability of food and feed. RIKILT is ISO 17025 and ISO 17043 accredited (the accredited tests are described on www.ra.nl (no. L014 and R012)).

Appendix 4. Copy of letter/instructions sent together with test samples (continued)

DATE
December 17, 2018

PAGE
2 of 2

- Please provide method information using the excel file:
'ICI-EQUAS study DINCH-round_02_Method information v1.xlsx'
This excel file will be send to you by email (there is no need to provide
method information and LOQs via the web application).

Please contact me if you have any questions or need any assistance.

With kind regards,

Ingrid Elbers
Pt.rikilt@wur.nl

Appendix 5. Scope and LOQs as provided in the method information submitted by the laboratories.

	OH-MINCH	cx-MINCH	
Lab code	LOQ (ng/ml)		total
PT9712	0.4	0.5	2
PT9713	0.2	0.2	2
PT9717	0.25	0.25	2
PT9720	0.37	0.37	2
PT9724	1	0.5	2
PT9725	0.2	0.2	2
PT9726	0.14	0.1	2
PT9727	0.2	nt	1
PT9728	0.1	0.1	2
PT9741	0.5	0.5	2
PT9753	0.1	0.1	2
PT9754	0.05	0.05	2
total	12	11	

nt: not tested/not in the scope of the method

Appendix 6. Assigned values and Z-scores



Control material	Test sample A		Test sample B		Test sample A		Test sample B	
Biomarker	OH-MINCH				cx-MINCH			
Conc. hom. test (ng/ml)	6.91		23.0		3.53		12.1	
Assigned value (ng/ml)	6.91		23.0		3.66		12.1	
Uncertainty	1.3%		0.9%		2.3%		0.1%	
Robust mean	u>>		u>>		u>>		u>>	
Study RSD _R	51%		49%		55%		70%	
Lab code	ng/ml	Z-score	ng/ml	Z-score	ng/ml	Z-score	ng/ml	Z-score
PT9712	7.07	0.1	23.3	0.1	3.82	0.2	12.0	0.0
PT9713	8.8	1.1	26.9	0.7	5.5	2.0	16.6	1.5
PT9717	2.39	-2.6	11.4	-2.0	0.58	-3.4	3.42	-2.9
PT9720	4.611	-1.3	14.105	-1.5	1.869	-2.0	7.108	-1.6
PT9724	5.5	-0.8	17.85	-0.9	2.66	-1.1	7.11	-1.6
PT9725	6.7	-0.1	23.7	0.1	4.2	0.6	12.7	0.2
PT9726	8.35	0.8	25.4	0.4	3.80	0.2	13.0	0.3
PT9727	9.7156	1.6	36.6741	2.4	nt		nt	
PT9728	16.16	5.4	4.06	-3.3	5.27	1.8	1.19	-3.6
PT9741	11.3	2.5	30.4	1.3	5.9	2.4	16.3	1.4
PT9753	6.744	-0.1	22.603	-0.1	3.679	0.0	12.074	0.0
PT9754	6.9	0.0	23.0	0.0	3.5	-0.1	12.0	0.0

Assigned value is the expert-assigned value (mean of concentrations of three expert labs).

Study RSD_R is based on results from participants (excluding results from expert labs)

u>> = uncertainty of robust mean based on participants' results too large to calculate a meaningful robust mean

Appendix 7. Graphical representation of the Z-scores



Appendix 7. Graphical representation of the Z-scores (continued)

