



science and policy
for a healthy future

HORIZON2020 Programme
Contract No. 733032 HBM4EU

REPORT OF THE WP9 ICI

Round 03/2019

Chromium in blood

Version / date of issue	1 / 08-08-2019
Organiser	Institute and Outpatient Clinic of Occupational, Social and Environmental Medicine (IPASUM) Friedrich-Alexander University of Erlangen-Nuremberg Henkestr. 9-11 91054 Erlangen GERMANY
Coordinator	Thomas Göen, thomas.goeen@fau.de
Authors (IPASUM)	Stefanie Nübler, Moritz Schäfer, Karin H. A. Zarrabi
Approved by:	Thomas Göen (IPASUM)

Table of contents

Table of contents	2
1 Summary	3
2 Introduction	4
2.1 Confidentiality	4
3 Control material	5
3.1 Preparation of control material	5
3.2 Homogeneity of control material	5
3.3 Stability of control material	5
4 Organisational details	6
4.1 Participants.....	6
4.2 Dispatch and instructions	6
4.3 Deviations from ICI SOPs	6
5 Data evaluation	7
5.1 False negatives and <LOQ	7
5.2 Assigned value	7
5.3 Target standard deviation	7
5.4 ICI standard deviation	7
5.5 Z-scores.....	7
6 Results and discussion	8
6.1 Results submitted by participants.....	8
6.2 Assigned values and (target) standard deviations.....	8
6.3 Assessment of laboratory performance.....	8
6.4 Conclusions and recommendations	8
7 References	10

Appendices

1 Homogeneity data.....	11
2 Stability data.....	13
3 Copy of letter of invitation.....	14
4 Copy of letter/instructions sent together with test samples.....	16
5 Method information.....	17
6 Assigned values and participant`s performance.....	19
7 Graphical representation of Z-scores.....	20
8 Results and LOQs of the shipped test samples analysed by the participants.....	22

1 Summary

Within the framework of the HBM4EU project, an Inter-Laboratory Comparison Investigation (ICI) was organized and conducted for the analysis of chromium in blood.

The study was performed from May 2019 to July 2019.

In total, 46 laboratories were invited for this third round, of which 21 laboratories from 17 countries registered.

The participation in this ICI was satisfactory; 20 out of 21 laboratories (95%) submitted their results.

In June 2019, two different test samples consisting of 3 mL blood spiked with chromium at two different concentrations (Cr_{low} , Cr_{high}), one of each concentration, were prepared, and one sample of each concentration was sent to the participating laboratories for analysis.

Homogeneity and stability assessment of the control materials confirmed that the materials were adequately homogeneous and stable.

Laboratory results were rated using Z-scores in accordance with ISO 13528 and ISO 17043. The standard deviation for proficiency assessment (also called target standard deviation) was set to FFP = 25%, as described in 5.3.

The evaluation of Cr_{low} showed that 95% of the results were satisfactory and 5% were questionable (Table 1). The results for Cr_{high} were all satisfactory.

Table 1 Overview of results for chromium in blood in 3rd ICI

Number of laboratories with respective results for chromium in blood in 3 rd ICI				
biomarker	assigned value	satisfactory ($ Z\text{-score} \leq 2$)	questionable ($2 < Z\text{-score} < 3$)	unsatisfactory ($ Z\text{-score} > 3$)
Cr_{low}	1.773 ng/mL	19 (95%)	1 (5%)	0
Cr_{high}	5.296 ng/mL	20 (100%)	0	0

WP9 ICI Report, Round 03/2019	Version: 1	Date: 08-08-2019	Page: 4
Chromium in blood, Round 3			

2 Introduction

Inter-Laboratory Comparison Investigations (ICI) and External Quality Assurance Schemes (EQUAS) are tools to assess the proficiency of laboratories, and the comparability and reliability of analytical methods. Participation in ICI / EQUAS forms an integral part of quality control, in addition to initial and ongoing in-house method validation.

This ICI study has been organised within the frame of HBM4EU as part of the Quality Assurance program for biomonitoring analyses. Within HBM4EU, participation in ICI/EQUAS exercises is mandatory for laboratories that will analyse HBM4EU samples.

This report describes the 3rd round of proficiency testing for chromium in blood, which was organised by the Institute and Outpatient Clinic of Occupational, Social and Environmental Medicine (IPASUM) at Friedrich-Alexander University of Erlangen-Nuremberg.

2.1 Confidentiality

In this report, the identity of the participants and the information provided by them is treated as confidential. However, lab codes of the participants will be disclosed to the HBM-QAU for performance assessment.

WP9 ICI Report, Round 03/2019	Version: 1	Date: 08-08-2019	Page: 5
Chromium in blood, Round 3			

3 Control material

3.1 Preparation of control material

For control material, surrogate material was used. It consists of bovine blood of German origin in EDTA solution and sodium azide. Animal health conditions are certified. The stock solution (Chromium ICP standard, ammonium dichromate in H₂O, 1000 mg/L, J.T.Baker) was diluted into two different concentrations and the addition to the native control material resulted in the intended concentration in control material (Cr_{low}, Cr_{high}). The two spiked control materials were aliquoted (3 mL each) into tubes with caps (57x15.3 mm, polypropylene, Sarstedt). The tubes were stored in a freezer (≤ -18 °C) until transportation. The two different concentrations (Cr_{low}, Cr_{high}) were measured using ICP-MS (see analysis method in **Appendix 5**).

3.2 Homogeneity of control material

Ten tubes of each concentration of the control material (Cr_{low}, Cr_{high}) were randomly selected from the freezer (≤ -18 °C). The thawed samples were re-homogenised by vortex shaking and analysed in duplicate using ICP-MS (analysis method see **Appendix 5**). The homogeneity was evaluated according to ISO 13528:2015, Fearn et al [2001] and Thompson [2000]. The results are presented in **Appendix 1**. The conclusion is that no outliers are detected, homogeneity is adequate and the method is suitable.

3.3 Stability of control material

On the day of preparation of the control materials, six randomly selected test samples of Cr_{low} and six randomly selected test samples of Cr_{high} were stored at -80 °C. The assumption is that under these conditions, the biomarker (Cr) is stable in blood. On the last day of the deadline for submission of results by the participants (July 23, 2019), six test samples of each level (stored at -80 °C) and six samples of each level (stored at -18 °C) were thawed and re-homogenised by vortex shaking. Next, all samples were analysed using ICP-MS (analysis method see **Appendix 5**). The stability was evaluated according to HBM4EU-SOP-QA-002 and using the Excel sheet "HBM4EU ICI-EQUAS stability test CM v1". The results are presented in **Appendix 2**. No consequential instabilities and no statistical differences were detected.

4 Organisational details

4.1 Participants

A list of 46 candidate laboratories from different countries eligible for the analysis of chromium had been compiled by the Work Package (WP) Task 9.2 leaders and made available to the institution organizing the respective ICI.

Invitation letters were sent by e-mail to all 46 candidate laboratories on June 3, 2019 (see **Appendix 3**). It was indicated that participation would be free of charge and that those who subscribed to the ICI would receive a kit containing the test materials needed for analysis. The condition for participation was that the test results had to be submitted within the stipulated deadline (April 30, 2019).

Twenty-one laboratories (46%) from 17 countries out of the 46 laboratories in the candidate list indicated their interest in participating in this ICI and sent their registration forms to IPASUM with their agreement to abide by the conditions for participation. These laboratories received an individual laboratory code to report their measurement results (see **Appendix 8**).

Twenty of the 21 potential participants (95%) performed the assays and submitted their results. All participants reported their results within the stipulated deadline (July 23, 2019).

4.2 Dispatch and instructions

Test materials were dispatched to the participants under ambient conditions on June 25, 2019. Each participant received two test samples spiked with the biomarker at two levels, one of each concentration. Each sample consisted of approximately 3 mL blood.

Moreover, a letter with instructions on sample handling (instruction letter, see **Appendix 4**), a sample receipt form to be sent back to IPASUM upon receipt of the test material as well as a result submission form and a method information form were sent to the participants by e-mail. The latter form was used to extract relevant information related to the analytical method used for quantification.

Participants were asked to perform a single analysis of each sample using the same procedure as will be used for analysis of samples in the frame of HMB4EU and to report results following the instructions given.

4.3 Deviations from ICI SOPs

For this 3rd round, the HMB4EU-QA-SOPs were followed. There were no deviations from these SOPs.

5 Data evaluation

5.1 False negatives and <LOQ

Classification of false positives and biomarkers reported as "<LOQ-value" or "not detected" (ND) is described in HBM4EU-SOP-QA-003. In this EQUAS there were no false positives and no ND. Therefore, no further description is given here.

5.2 Assigned value

For ICI studies, the consensus value is used as assigned value and calculated as described in SOP HBM4EU-SOP-QA-003. In brief, the consensus value and its uncertainty were calculated from the results submitted by the participants using robust statistics to minimize the influence of outliers.

5.3 Target standard deviation

For calculation of the Z-scores, a fit-for-purpose relative target standard deviation (FFP-RSD_R) of 25% of the assigned value was used as target standard deviation. This was the default indicated in HBM4EU-SOP-QA-003 and considered appropriate based on the outcome of the 1st round.

5.4 ICI standard deviation

To gain insight into the actual inter-laboratory variability of the biomarker analysis in this study, the robust relative standard deviation (RSD_R) was calculated based on the participants' results, as described in HBM4EU-SOP-QA-003.

5.5 Z-scores

Z-scores were calculated according to SOP HBM4EU-SOP-QA-003.

In accordance with ISO 13528 and ISO 17043 and the deliverable D 9.4 "*The Quality Assurance/Quality Control Scheme in the HBM4EU project*", Z-scores are classified as presented in **Table 2**.

Table 2 Classification of Z-scores

$ Z \leq 2$	satisfactory
$2 < Z < 3$	questionable
$ Z \geq 3$	unsatisfactory

WP9 ICI Report, Round 03/2019	Version: 1	Date: 08-08-2019	Page: 8
Chromium in blood, Round 3			

6 Results and discussion

6.1 Results submitted by participants

In total, 21 laboratories from 17 countries agreed to participate in this study. Not all participants were able to meet the stipulated deadline due to technical problems, so that in the end, 20 out of 21 participants (95%) submitted their results. Laboratories were also asked to provide LOQs.

Appendix 8 gives an overview of results and LOQs submitted by the participants as well as reasons for delayed submission.

Results indicated as 'not detected' (ND, see Appendix 8):

For **Cr_{low}** and **Cr_{high}**, no laboratory indicated ND.

False positive results: No participant detected a false positive in **Cr_{low}** and **Cr_{high}**.

Methods:

In almost all cases, the samples were analysed by ICP-MS followed by AAS. For sample preparation, the majority of participating laboratories used no digestion. Almost all participating laboratories used a single quad as detection system, the rest used a triple quad or quadropol. As internal standard, most laboratories applied rhodium, germanium or yttrium and for calibration, most candidates used an external calibrant (matrix-based), followed by standard addition and external calibrant (solvent-based).

6.2 Assigned values and (target) standard deviations

The assigned value and its uncertainty (u), the robust relative standard deviation of the participant's data (study RSD_R), and the fit-for-purpose (FFP) target standard deviation for each of the control materials are included in Appendix 6.

6.3 Assessment of laboratory performance

Z-scores were calculated and the graphical representations of the Z-scores for **Cr_{low}** and **Cr_{high}** are provided in **Appendix 7**.

Twenty laboratories out of 21 registered candidate laboratories reported results. 19 laboratories (95%) have satisfactory Z-scores in **Cr_{low}** and in 20 laboratories (100%) achieved satisfactory Z-scores in **Cr_{high}**.

6.4 Conclusions and recommendations

The overall participation in the HBM4EU ICI round 3 was successful. Twenty-one laboratories out of the 46 laboratories (46%) in the candidate list confirmed their participation in the ICI. This ICI was specifically aimed at participants who want to qualify for the occupational chromium study. Consequently, successful participation in two ICI studies will lead to approval and a certificate being issued for the occupational exposure range only.

Twenty of these 21 registered candidate laboratories reported results, representing a participation rate of 95%.

Regarding the quantification of chromium in blood, the participants achieved in **Cr_{low}** 95% satisfactory and 5% questionable results. In **Cr_{high}**, all participants achieved 100% satisfactory results.

The participants with unsatisfactory or questionable results are recommended to do a root cause analysis to find the reason for the deviating results, and seek assistance from HBM4EU expert laboratories if needed.

A direct comparison of the overall performance of the laboratories with that of the first and second round is not entirely possible, because there were a lot of new laboratories participating in this round and also some laboratories from the first and second round did not participate in this third round. Moreover, in the first round, only one level could be evaluated for the participants and in the second and third round, two levels could be evaluated.

Two additional laboratories reported results in this third round compared to the second ICI round. The percentage of satisfactory Z-scores at the low and the high level was 85% in the second round, 95% and 100% in this third round.

Table 3 below gives an overview of the performance of the individual laboratories in this round for chromium in blood.

Table 3 Performance of the candidate laboratories for chromium in blood

Lab code	LOQ [ng/mL]	Cr _{low}	Cr _{high}
QR/104	1.800	satisfactory	satisfactory
QR/105	2.500	satisfactory	satisfactory
QR/107	0.028	satisfactory	satisfactory
QR/110	0.500	satisfactory	satisfactory
QR/111	0.220	satisfactory	satisfactory
QR/112	0.100	satisfactory	satisfactory
QR/113	0.100	satisfactory	satisfactory
QR/115	0.600	satisfactory	satisfactory
QR/129	0.110	satisfactory	satisfactory
QR/130	0.500	satisfactory	satisfactory
QR/143	0.360	satisfactory	satisfactory
QR/202	0.200	satisfactory	satisfactory
QR/204	0.500	satisfactory	satisfactory
QR/205	0.050	satisfactory	satisfactory
QR/207	0.344	satisfactory	satisfactory
QR/211	0.100	satisfactory	satisfactory
QR/212	0.100	satisfactory	satisfactory
QR/224	0.050	questionable	satisfactory
QR/225	0.050	satisfactory	satisfactory
QR/301	0.200	satisfactory	satisfactory

7 References

- [1] Analytical Methods Committee, 1989a, Robust statistics - How not to reject outliers Part 1. Basic concepts, Analyst, 114, 1693-1697.
- [2] Analytical Methods Committee, 1989b, Robust statistics - How not to reject outliers Part 2. Interlaboratory trials, Analyst, 114, 1699-1702
- [3] HBM4EU-SOP-QA-001 "Organisation of Interlaboratory Comparison Investigations (ICI) and External Quality Assurance Schemes (EQUAS) of interlaboratory studies"
- [4] HBM4EU-SOP-QA-002 "Preparation of test materials for ICI / EQUAS"
- [5] HBM4EU-SOP-QA-003 "Evaluation of ICI / EQUAS results"
- [6] HBM4EU-SOP-QA-004 "Reporting of ICI / EQUAS studies"
- [7] ISO/IEC 17043:2010, Conformity assessment – General requirements for proficiency testing
- [8] ISO 13528, 2015, Statistical methods for use in proficiency testing by interlaboratory comparison.
- [9] Official Methods of Analysis Program Manual, 2002, Appendix D: Guidelines for Collaborative Study Procedures to Validate Characteristics of a Method of Analysis. Association of Analytical Communities International.
http://www.aoac.org/vmeth/Manual_Part_6.pdf.
- [10] Thompson, M., 2000, Recent trends in inter-laboratory precision at ppb and sub-ppb concentrations in relation to fitness for purpose criteria in proficiency testing, Analyst, 125, 385-386.
- [11] Thompson M., Ellison R. and Wood, R., 2006, The International Harmonized Protocol for the Proficiency Testing of Analytical Chemistry Laboratories, Pure Appl. Chem, 78(1), 145-196.

Appendix 1. Homogeneity data

Homogeneity

Version HBM4EU v1

Control material: **blood**

Analyte: **Chromium**

Preparation of control material: **low**
10 randomly chosen test samples, analysed in duplicate

Target standard deviation:

Fit-for-purpose RSD **FFP** (25% is default value)
if you want to use Horwitz/Thompson,
then delete FFP from cell H5

[1] ISO 13528:2005

[2] Fearn, T. and M. Thompson, 2001, A New Test for 'Sufficient homogeneity', Analyst, 126, 1414-1417

[3] Thompson M., 2000, Recent trends in inter-laboratory precision at ppb and sub-ppb concentrations in relation to fitness for purpose criteria in proficiency testing, Analyst, 125, 385-386

	replicate 1	replicate 2	x_i	w_i	w_i^2	$(x_i - \bar{x})^2$
1	2.259	1.887	2.073	0.372	0.138	0.006
2	1.993	2.042	2.018	-0.049	0.002	0.001
3	1.957	1.879	1.918	0.078	0.006	0.006
4	2.009	1.976	1.993	0.033	0.001	0.000
5	2.098	1.967	2.033	0.131	0.017	0.002
6	2.140	1.888	2.014	0.252	0.064	0.000
7	1.982	2.094	2.038	-0.112	0.013	0.002
8	1.962	2.155	2.059	-0.193	0.037	0.004
9	1.956	1.828	1.892	0.128	0.016	0.010
10	1.938	1.852	1.895	0.086	0.007	0.010

Lowest: 1.828 µg/kg $\Sigma =$ 0.302 0.041
Highest: 2.259 µg/kg
Grand mean (\bar{x}): 1.993 µg/kg
Stdev: 0.111 µg/kg
VC%: 5.6% µg/kg

Outliers: Cochran's test

$$C = w_{\max}^2 / \Sigma w_i^2$$

→ C = 0.458

→ Ccrit = 0.602 **C < Ccrit → No outliers detected**

Horwitz [3]:

Mean > 120 ppb: $CV = 2(1 - \frac{1}{2} \log c)$

Mean < 120 ppb: $\sigma = 0.22c$

FFP (fit-for-purpose)

RSD% = 40.79

RSD% = 22

RSD% = 25

σ_H = 0.813

σ_H = 0.438

σ_H = 0.498

σ_H used: 0.498

Homogeneity [1]:

s_x = 0.067

s_w = 0.123 (within sample standard deviation)

s_s = 0.000 (between sample standard deviation)

critical = 0.149

$s_s < \text{critical?}$ → **ACCEPT: Homogeneity adequate**

$s_w < 0.5 \cdot \sigma_H?$ → **ACCEPT: Method suited**

Appendix 1 Homogeneity data (continued)

Homogeneity

Version HBM4EU v1

Control material: **blood**

Analyte: **Chromium**

Preparation of control material: **high**

10 randomly chosen test samples, analysed in duplicate

Target standard deviation:

Fit-for-purpose RSD **FFP** (25% is default value)
if you want to use Horwitz/Thompson,
then delete FFP from cell H5

[1] ISO 13528:2005

[2] Fearn, T. and M. Thompson, 2001, A New Test for 'Sufficient homogeneity', Analyst, 126, 1414-1417

[3] Thompson M., 2000, Recent trends in inter-laboratory precision at ppb and sub-ppb concentrations in relation to fitness for purpose criteria in proficiency testing, Analyst, 125, 385-386

	replicate 1	replicate 2	x_i	w_i	w_i^2	$(x_i - \bar{x})^2$
1	5.540	5.206	5.373	0.334	0.112	0.002
2	5.086	5.189	5.138	-0.103	0.011	0.038
3	4.952	5.144	5.048	-0.192	0.037	0.080
4	5.174	5.161	5.168	0.013	0.000	0.027
5	5.932	5.062	5.497	0.870	0.757	0.027
6	5.416	5.427	5.422	-0.011	0.000	0.008
7	5.188	5.182	5.185	0.006	0.000	0.021
8	5.318	5.673	5.496	-0.355	0.126	0.027
9	5.101	5.517	5.309	-0.416	0.173	0.001
10	5.825	5.537	5.681	0.288	0.083	0.122

Lowest: 4.952 µg/kg $\Sigma =$ 1.298 0.353
Highest: 5.932 µg/kg
Grand mean \bar{x} : 5.332 µg/kg
Stdev: 0.267 µg/kg
VC%: 5.0% µg/kg

Outliers: Cochran's test

$$C = w_{\max}^2 / \Sigma w_i^2$$

→ C = 0.583

→ Ccrit = 0.602 **C < Ccrit → No outliers detected**

Horwitz [3]:

Mean > 120 ppb: CV=2(1-1/2 log c)

Mean < 120 ppb: $\sigma = 0.22c$

FFP (fit-for-purpose)

RSD% = 35.18

RSD% = 22

RSD% = 25

σ_H = 1.875

σ_H = 1.173

σ_H = 1.333

σ_H used: 1.333

Homogeneity [1]:

s_x = 0.198

s_w = 0.255 (within sample standard deviation)

s_s = 0.082 (between sample standard deviation)

critical = 0.400

$s_s < \text{critical?}$ → **ACCEPT: Homogeneity adequate**

$s_w < 0.5 \cdot \sigma_H$? → **ACCEPT: Method suited**

Appendix 2 Stability data

HBM4EU ICI-EQUAS stability test CM v1

Stability test in frame of ICI-EQUAS

Material: **Cr (blood)**
Storage: minus 80 °C minus 18 °C

Biomarker **Chromium low**

	t=0 (storage)	t=a (analysis)
dates:	05.08.2019	05.08.2019
values:	1.59	1.96
	1.82	1.70
	2.01	1.75
	1.84	1.68
	2.12	1.69
	1.94	2.05
number=	6	6
average=	1.89	1.81
std dev=	0.182	0.159

analysis date	time (days)	µg/L	n	std dev
05.08.2019	0	1.89	6	0.182
05.08.2019	0	1.81	6	0.159
x0-xa=		0.08		

test 'consequential instability'	Horwitz/Thompson	Fit-for-purpose (FFP)
xav-yav =<0,3σH		
σH=	0.415	0.4715
0,3*σH=	0.124	0.14145
x0-xa<0,3*σH?	No consequential instability detected	No consequential instability detected

test 'significant difference':	
F=	1.32
Fcrit=	5.05
Significant difference?	No significant difference in std detected
sed^2=	0.17
n=	10
std difference=	0.10
t=	0.82
t-crit=	2.23
Significant difference?	No statistic instability detected

HBM4EU ICI-EQUAS stability test CM v1

Stability test in frame of ICI-EQUAS

Material: **Cr (blood)**
Storage: minus 80 °C minus 18 °C

Biomarker **Chromium high**

	t=0 (opslag)	t=a (analyse)
dates:	05.08.2019	05.08.2019
values:	5.89	4.84
	5.63	5.90
	5.13	5.70
	4.89	5.20
	4.83	5.14
	5.50	4.99
number=	6	6
average=	5.31	5.29
std dev=	0.428	0.416

analysis date	time (days)	µg/L	n	std dev
05.08.2019	0	5.31	6	0.428
05.08.2019	0	5.29	6	0.416
x0-xa=		0.02		

test 'consequential instability'	Horwitz/Thompson	Fit-for-purpose (FFP)
xgem-ygem =<0,3σH		
σH=	1.169	1.32825
0,3*σH=	0.351	0.398475
x0-xa<0,3*σH?	No consequential instability detected	No consequential instability detected

test 'significant difference':	
F=	1.06
Fcrit=	5.05
Significant difference?	No significant difference in std detected
sed^2=	0.42
n=	10
std difference=	0.24
t=	0.08
t-crit=	2.23
Significant difference?	No statistic instability detected

WP9 ICI Report, Round 03/2019	Version: 1	Date of issue: 08-08-2019	Page: 14
Chromium in blood, Round 3			

Appendix 3 Copy of letter of invitation

HBM4EU: Announcement / invitation to participate in ICI study Cr/Round 3

Title of ICI: Chromium in blood, serum and urine

Dear Colleagues,

within the frame of HBM4EU the

Institute and Outpatient Clinic of Occupational, Social and Environmental Medicine
Friedrich-Alexander University Erlangen-Nuremberg
Henkestr. 9-11
91054 Erlangen
Germany

announces the 3rd ICI round for the determination of chromium in blood, serum and urine. The aim of ICI exercises is to provide laboratories with an assessment of their analytical performance and reliability of their data in comparison with other laboratories. This will aid in the quality improvement of analysis in human biomonitoring at each of the laboratories.

Test samples

The matrices will be blood, serum and urine, respectively. Accordingly, the participants will receive:

- 2 different materials of blood (1 sample of 3 mL each) for determination of chromium in blood and/or
- 2 different materials of serum (1 sample of 3 mL each) for determination of chromium in serum and/or
- 2 different materials of urine (1 sample of 5 mL each) for determination of chromium in urine

Target biomarkers

The biomarker potentially present in the test samples is chromium.

This ICI is specifically aimed at participants who want to qualify for the occupational chromium study. Consequently, successful participation in the ICI studies will lead to approval and a certificate being issued for the occupational exposure range only.

With regard to the conditions and specifications of the occupational study, we expect participants of this ICI to use suitable analytical methods to obtain an **LOQ < 1 µg/L**.

Participation in this ICI is mandatory for laboratories that want to get approved for the occupational chromium study, irrespective of their participation in the first two ICIs.

WP9 ICI Report, Round 03/2019	Version: 1	Date: 08-08-2019	Page: 15
Chromium in blood, Round 3			

Calendar:

Registration deadline	June 17, 2019 (we would be pleased to receive your feedback as soon as possible)
Distribution of test samples (projected)	June 25, 2019
Deadline for submission of results (projected)	July 23, 2019

Registration

For registration, please find attached a registration form each for chromium in blood, in serum and in urine. Please send them back to us by mail in case you want to register.

Upon registration, the participant will receive a lab-code to be used for submission of results.

Fee

For partners and linked-third parties of HBM4EU, participation is free of charge. Please note that the participant is responsible for custom clearance and associated costs if applicable.

Confidentiality:

All laboratory-specific information will be treated confidentially, and will never be disclosed to third parties (government, accreditation bodies) except the HBM4EU QAU, without permission of the laboratory

Contact information organiser:

Coordinators:

- Prof. Dr. Thomas Göen
- Stefanie Nübler
- Moritz Schäfer
- Karin H. A. Zarrabi

Institute and Outpatient Clinic of Occupational, Social and Environmental Medicine
Friedrich-Alexander University Erlangen-Nuremberg
Henkestr. 9-11
91054 Erlangen
Germany

Email: ipasum-hbm4eu@fau.de

WP9 ICI Report, Round 03/2019	Version: 1	Date: 08-08-2019	Page: 16
Chromium in blood, Round 3			

Appendix 4 Copy of letter/instructions sent together with test samples

HBM4EU: Instruction letter ICI study Cr in blood/Round 3

Title of ICI: Chromium in blood

Dear participant,

Thank you for participation in the HBM4EU ICI study Cr in blood/Round 3 for the determination of Chromium in blood.

You will receive a parcel containing 2 test samples spiked with the biomarker at 2 levels, 1 of each concentration. Each sample consists of approximately 3 mL blood.

The parcel will be shipped on 25 June 2019 under ambient conditions.

Instructions:

- Upon receipt, please check the content for any damage/leaking of the containers, complete the sample receipt form and return it to the organiser.
- Store the test samples under frozen (-18°C) conditions until analysis.
- Analyse the samples for the biomarkers indicated in the invitation letter ref/ 03.06.2019.
- Thaw the samples and re-homogenise them according to your own procedure.
- Analyse the samples using the same procedure as will be used for analysis of samples in the frame of HBM4EU.
- Carry out a single analysis for each sample.
- For submission of results and method information, use the forms provided.
- The deadline for submission of analysis results and method details is **23 July 2019**.

If you have any questions or need any assistance, please contact:

Karin Zarrabi, Stefanie Nübler or Moritz Schäfer

Email: ipasum-hbm4eu@fau.de

Tel.: + 49 (0)9131/8526146, /8526145

Institute and Outpatient Clinic for Occupational, Social and Environmental Medicine
Friedrich-Alexander University Erlangen-Nuremberg
Henkestrasse 9-11
91054 Erlangen
Germany

Prof. Dr. Thomas Göen (for the ICI organisers)

WP9 ICI Report, Round 03/2019	Version: 1	Date: 08-08-2019	Page: 17
Chromium in blood, Round 3			

Appendix 5 ICP-MS method information IPASUM

HBM4EU: Method information form for participation in ICI / EQUAS study Cr/Round 3


Title of ICI: Chromium in blood

Laboratory code	IPASUM Erlangen - Germany
ISO17025 accredited	no
<u>SAMPLE PREPARATION</u>	
amount sample	0.2 mL
<u>Extraction</u>	
- pH adjustment	no
- LLE; solvent(s) / time / shaking	
- SPE; material	
<u>Digestion</u>	
	no
<u>INSTRUMENTAL ANALYSIS</u>	
AAS	no
Wavelength	
Background compensation	
Matrix modifier	
Dilution factor	
Other remarks	
<u>ICP</u>	
	Yes (ICP-MS Perkin Elmer)
Dilution	1:20 with H ₂ O, NH ₄ , EDTA and Triton-X-solution
Nebulizer	glass
Reagent gas	Argon
Masses monitored	Cr ⁵²
<u>Detection</u>	
MS	single quad
OES	
<u>Quantification</u>	
Use of internal standard (IS)	yes (Rh ¹⁰³)
- response normalised to IS	yes
<u>Calibration</u>	
	external calibrant (matrix-based)
	Multi-level
Correction for recovery	no
<u>Identification criteria used</u>	
- ion ratio tolerance	% relative
- other	-


Appendix 6 Assigned values and participant's performance

HBM4EU 03/2019	Cr (blood)			
control material	Cr_{low}		Cr_{high}	
assigned value	1.773 ng/mL		5.296 ng/mL	
uncertainty of assigned value	0.090 ng/mL		0.154 ng/mL	
study RSD_R	18.1%		10.4%	
relative target standard deviation	25%		25%	
laboratory code	value	Z-score	value	Z-score
QR/104	2.005	0.52	4.546	-0.57
QR/105	2.650	1.98	6.220	0.70
QR/107	1.607	-0.37	4.911	-0.29
QR/110	1.322	-1.02	5.067	-0.17
QR/111	1.641	-0.30	4.394	-0.68
QR/112	1.500	-0.61	5.320	0.02
QR/113	1.760	-0.03	5.470	0.13
QR/115	1.270	-1.13	4.750	-0.41
QR/129	1.460	-0.71	4.920	-0.28
QR/130	1.820	0.11	5.670	0.28
QR/143	1.900	0.29	5.520	0.17
QR/202	1.650	-0.28	5.570	0.21
QR/204	1.929	0.35	5.720	0.32
QR/205	1.923	0.34	4.912	-0.29
QR/207	1.675	-0.22	5.181	-0.09
QR/211	1.290	-1.09	4.470	-0.62
QR/212	2.000	0.51	5.430	0.10
QR/224	2.746	2.20	6.256	0.72
QR/225	2.201	0.97	5.729	0.33
QR/301	1.854	0.18	5.732	0.33

Appendix 7 Graphical representation of the Z-scores



Appendix 7 Graphical representation of the Z-scores (continued)



Appendix 8 Results and LOQs and reasons for delayed submission

HBM4EU 3-2019 Chromium in blood [ng/mL]				
Lab.code	low	high	LOQ	delayed reporting
QR/104	2.005	4.546	1.800	
QR/105	2.650	6.220	2.500	
QR/107	1.607	4.911	0.028	
QR/110	1.322	5.067	0.500	
QR/111	1.641	4.394	0.220	
QR/112	1.500	5.320	0.100	
QR/113	1.760	5.470	0.100	
QR/115	1.270	4.750	0.600	
QR/129	1.460	4.920	0.110	
QR/130	1.820	5.670	0.500	
QR/143	1.900	5.520	0.360	
QR/202	1.650	5.570	0.200	
QR/204	1.929	5.720	0.500	
QR/205	1.923	4.912	0.050	
QR/207	1.675	5.181	0.344	
QR/211	1.290	4.470	0.100	
QR/212	2.000	5.430	0.100	
QR/224	2.746	6.256	0.050	
QR/225	2.201	5.729	0.050	
QR/301	1.854	5.732	0.200	