

Report of the ICI/EQUAS

Round 01/2019

Anilines (aromatic amines) in urine

Version/ date of issue	1.0/ 09/20/2019
Organizer	Analytisch-Biologisches Forschungslabor (ABF) GmbH Simmelweisstr. 5 82152 Planegg Germany
Coordinator	Dr. Nikola Pluym, nikola.pluym@abf-lab.com Dr. Max Scherer, max.scherer@abf-lab.com
Authors	Dr. Therese Burkhardt, Dr. Nikola Pluym
Approved by	Thomas Göen (Institute and Outpatient Clinic of Occupational, Social and Environmental Medicine)

Table of contents

1	Summary	2
2	Introduction	4
3	Confidentiality	4
4	Control materials	5
4.1	Preparation of control material	5
4.2	Homogeneity of control material.....	5
4.3	Stability of control material	5
5	Organizational details	6
6	Dispatch and instructions	6
7	Data evaluation.....	6
7.1	False positives and <LOQ	6
7.2	Assigned value	7
7.3	Target standard deviation (σ_T)	8
7.4	Z-scores	8
8	Results and Discussion	8
8.1	Participants' Results.....	8
8.2	Assigned values and (target) standard deviations	8
8.3	Assessment of laboratory performance	9
8.4	Conclusion and recommendations	9
9	References	12
	Appendix	I
I	Homogeneity data	II
II	Stability data.....	IV
III	Method information form for the participation in EQUAS/ICI (Round 1)	VI
I.I	Method information for measuring homogeneity and stability of AN and TOL	VI
I.II	Method information for measuring homogeneity and stability of 2,4-TDA, 2,6-TDA, MDA, and MOCA	VIII
IV	Copy of invitation letter	X
V	Copy of registration form	XII
VI	Copy of instruction letter	XIV
VII	Results of the control material analyzed by the participants	XV
VIII	Assigned values and participants' performance for Anilines in urine	XVII
IX	Graphical evaluation of Z-Scores	XX

1 Summary

Within the framework of the HBM4EU project, an External Quality Assurance Scheme (EQUAS) and an Inter-laboratory Comparison Investigation (ICI) was organized and conducted for the analysis of aromatic amines (AA) in urine.

The study was performed from May 2019 until July 2019. In total, 18 laboratories were invited for this 1st EQUAS/ICI Round, of which nine laboratories from four countries registered. Eight laboratories submitted results yielding in a participation rate of 89 %.

In May 2019, six different test samples of low and high concentration, consisting of 8 mL urine spiked with aromatic amines were sent on dry ice to the participating expert laboratories for single analysis. Candidate laboratories received six samples, three for the low and high concentration levels, respectively, for single analysis. The HBM4EU QAU selected four expert laboratories for analysis of aromatic amines in urine (at least three per parameter).

Assessment of the control material according to Thompson (2006) confirmed the adequate homogeneity for all substances except for 4,4'-methylenebis(2-chloraniline) (MOCA). However, with additional allowance for sampling errors and repeatability, homogeneity for MOCA was sufficient (Fearn and Thompson, 2001). No significant instability was detected for the aromatic amines investigated, except for *ortho*-toluidine (TOL) in both levels. The decrease in concentration over time was considered when calculating Z-Scores.

The performance of the laboratories was assessed by calculating Z-Scores using the assigned value, mean of expert laboratories (EQUAS) or robust mean of all participants (ICI), and a fixed fit-for-purpose target standard deviation of 25 %. Assigned values and thus Z-scores could be calculated for all aromatic amines except for aniline (AN), since the requirements for evaluating the Z-Scores were not met. An overview of the results for the analytes investigated is given in Table 1.

Table 1: Overview of the results for aromatic amines in urine, round 1

Analyte ¹	participants	evaluation scheme	expert labs (after exclusion of outliers)	assigned value	Satisfactory	Questionable	Unsatisfactory
2,4-TDA _{Low}	7	EQUAS	3	49.1 ng/mL	5 (71 %)	1 (14 %)	1 (14 %)
2,4-TDA _{High}	7	EQUAS	3	178.3 ng/mL	5 (71 %)	1 (14 %)	1 (14 %)
2,6-TDA _{Low}	7	EQUAS	3	51.8 ng/mL	5 (71 %)	0 (0 %)	2 (29 %)
2,6-TDA _{High}	7	ICI	-	196.2 ng/mL	5 (71 %)	1 (14 %)	1 (14 %)
AN _{Low}	4	-	-	not evaluated	no Z-scores	no Z-scores	no Z-scores
AN _{High}	4	-	-	not evaluated	no Z-scores	no Z-scores	no Z-scores
MDA _{Low}	8	EQUAS	4	4.9 ng/mL	8 (100 %)	0 (0 %)	0 (0 %)
MDA _{High}	8	EQUAS	2	95.2 ng/mL	7 (88 %)	1 (13 %)	0 (0 %)
MOCA _{Low}	8	ICI	-	8.5 ng/mL	6 (75 %)	1 (13 %)	1 (13 %)
MOCA _{High}	8	EQUAS	3	107.6 ng/mL	5 (63 %)	2 (25 %)	1 (13 %)
TOL _{Low}	4	EQUAS	3	0.32 ng/mL	4 (100 %)	0 (0 %)	0 (0 %)
TOL _{High}	4	EQUAS	3	1.4 ng/mL	4 (100 %)	0 (0 %)	0 (0 %)

¹ Abbr.: 2,4-TDA: 2,4-diaminotoluene; 2,6-TDA: 2,6-diaminotoluene; AN: aniline; MDA: 4,4'-methylenedianiline; MOCA: 4,4'-methylenebis(2-chloraniline); TOL: *ortho*-toluidine.

For 2,6-TDA_{High} and MOCA_{Low}, the assigned value was determined using the consensus value of the ICI (SOP HBM4EU-SOP-QA-003). Table 2 gives an overview the laboratories performance for each analyte, both for the low and high concentration level.

Table 2: Results of participating laboratories for corresponding aromatic amines in urine in the low and high concentration range with (+) satisfactory, (o) questionable, (-) unsatisfactory, (NA) not analyzed, (NE) not evaluable.

Lab	2,4-TDA		2,6-TDA		AN		MDA		MOCA		TOL		SUM of biomarker analyzed
	Low	High	Low	High	Low	High	Low	High	Low	High	Low	High	
AA_01	+	+	+	+	NE	NE	+	+	o	+	+	+	5
AA_03	-	-	-	-	NA	NA	+	+	+	o	NA	NA	4
AA_05	NA	NA	NA	NA	NE	NE	NA	NA	NA	NA	NA	NA	0
AA_06	NA	NA	NA	NA	NA	NA	+	+	+	+	NA	NA	2
AA_07	+	+	+	+	NA	NA	+	+	+	+	NA	NA	4
AA_10	+	+	+	+	NA	NA	+	+	+	o	+	+	5
AA_12	+	+	+	+	NA	NA	+	+	+	+	NA	NA	4
AA_16	o	o	-	o	NE	NE	+	o	-	-	+	+	5
AA_21	+	+	+	+	NE	NE	+	+	+	+	+	+	5
SUM	7	7	7	7	0	0	8	8	8	8	4	4	

2 Introduction

Inter-Laboratory Comparison Investigations (ICI) and External Quality Assurance Schemes (EQUAS) are tools to assess the proficiency of laboratories, and the comparability and reliability of analytical methods. Participation in ICI / EQUAS forms an integral part of quality control, in addition to initial and on-going in-house method validation.

This ICI/EQUAS study has been organized within the frame of HBM4EU as part of the Quality Assurance program for biomonitoring analyses. Within HBM4EU, participation in ICI/EQUAS exercises is mandatory for laboratories that will analyze HBM4EU samples.

This report describes the 1st round for anilines (aromatic amines, AA) in urine and was organized by ABF GmbH. Six anilines were included in this 1st round of the EQUAS/ICI (Table 3). EQUAS assessment requires at least three expert laboratories (HBM4EU-SOP-QA-001) to calculate the expert assigned value by averaging the values obtained by the expert laboratories. If EQUAS evaluation was not possible, the consensus value – the robust mean of all participants – was used as assigned value.

Table 3: Aromatic amines analyzed in this EQUAS/ICI Round 1

Aromatic amines	Abbreviation
2,4-Diaminotoluene	2,4-TDA
2,6-Diaminotoluene	2,6-TDA
Aniline	AN
4,4'-Methylenedianiline	MDA
4,4'-Methylenebis(2-chloroaniline)	MOCA
o-Toluidine	TOL

3 Confidentiality

In this report, the identity of the participants and the information provided by them is treated confidential. However, lab codes of the participants will be disclosed to the HBM-QAU for performance assessments.

4 Control materials

4.1 Preparation of control material

For control material two different pools of non-smoker urine were adjusted to pH 4.0 and spiked with all aromatic amines investigated for the high concentration level (AA_{High}) and all aromatic amines except AN for the low concentration level (AA_{Low}).

Humane urine was spiked with a stock solution mixture containing aniline (AN), o-toluidine (TOL), 2,4-diaminotoluene (2,4-TDA), 2,6-diaminotoluene (2,6-TDA), 4,4'-methylenedianiline (MDA), an 4,4'-methylenebis(2-chloroaniline) (MOCA). Control material was aliquoted with a volume of 8 mL in 15 mL falcon tubes (PP, Greiner, Germany) and stored at -20 °C until shipment. Analytical methods used for measurement of homogeneity and stability are described in Appendix III. Measured concentrations are given in Appendix I (Table 6) and Appendix II (Table 7) of this report.

4.2 Homogeneity of control material

For determining the homogeneity, ten tubes per level of each control material (AA_{Low} and AA_{High}) were randomly selected from the freezer (-20 °C). Samples were thawed, homogenized by vortex shaking and analyzed in duplicates using the corresponding method (see Appendix III).

Homogeneity was evaluated according to HBM4EU-SOP-QA-002, Fearn (2001) and Thompson (2006) (Fearn and Thompson 2001, Thompson et al. 2006). Results of the homogeneity testing are given in Appendix I. No outliers were detected. Except for MOCA (low and high) and AN (low) the statistical analysis method used is suited to determine inhomogeneity. Homogeneity according to Thompson (2006) was adequate for all analytes except for MOCA. With additional allowance for sampling errors and repeatability, homogeneity was sufficient for MOCA with the method described by Fearn and Thompson (2001).

4.3 Stability of control material

Stability testing was performed according to HBM4EU-SOP-QA-002. Therefore, three samples each were randomly selected from the freezer (-20 °C) and analyzed at $t = 0$ d (day of control material preparation) and $t \geq 44$ d. Control samples (AA_{Low} and AA_{High}) were analyzed by ABF with the methods described in Appendix III. Assessment of the stability was done by comparing the means of the triplicates. The stability was evaluated according to HBM4EU-SOP-QA-002 and using the Excel sheet 'HBM4EU ICI-EQUAS stability test CM v1'. The results are presented in Appendix II. Consequential instability was detected for 2,6-TDA_{Low}, AN_{Low}, MDA_{Low}, MOCA_{Low}, and TOL_{High}. However, except for TOL, no significant difference ($\alpha = 0.05$) was observed within the investigated time period. Significant instability of TOL was taken into account when calculating Z-scores.

5 Organizational details

For the organization of the 1st ICI/EQUAS, the HBM4EU consortium conducted a survey to find expert laboratories for the determination of anilines in urine. Four expert laboratories, all from Europe, were selected according to HBM4EU-SOP-QA-001 and in agreement with HBM4EU QAU.

ABF sent the invitation letters (see Appendix III) and a registration form (Appendix V) by e-mail on December 10, 2018. It was indicated that participation would be free of charge and that the participants receive test materials needed for analysis.

Nine laboratories signed in for the participation giving their agreement to abide by the conditions for participation. Each laboratory received an individual lab code in order to report their measurement results. All participating laboratories including the expert labs had to submit test results within the stipulated deadline (July 5, 2019). Eight of nine laboratories performed the assays and submitted within the deadline agreed.

6 Dispatch and instructions

Control material was dispatched on dry ice to the participants on May 27, 2019. The expert laboratories received twelve, the other participants six samples (8 mL each) of two different levels of the test material. Moreover, instructions on sample handling (Appendix VI), a sample receipt form to be sent back to ABF as well as a result submission and a method information form (example see Appendix III) were sent to the participants by e-mail. The latter form was used to obtain relevant information on the analytical method used for quantification.

7 Data evaluation

7.1 False positives and <LOQ

Classification of false positives and biomarkers reported as "<LOQ value" or "not detected" (ND) was as described in HBM4EU-SOP-QA-003.

A result is assigned as false positive, if (1) the reported concentration is below the LOQ of the organizer, the expert laboratories, and the majority of the participants and (2) a concentration for a biomarker is reported that was not present in the control material.

If a biomarker is reported as '<LOQ' and an assigned value for the biomarker can be determined in the control material, further assessment is done to verify whether this result might be a false negative and the LOQ is considered adequate (low enough) for analysis within the frame of HBM4EU.

Results are classified as false negative, when no numerical value is reported for a biomarker present in the test material, although the laboratories LOQ is well below the assigned value.

LOQ is considered insufficient (too high), if (1) the LOQ is well above the assigned value, (2) the assigned value represents a realistic concentration of real samples within the framework of HBM4EU, and (3) quantitative determination is feasible by the majority of laboratories.

In order to judge results <LOQ in a quantitative way, 'proxy-Z-scores' need to be calculated. However, since no results <LOQ were reported in this EQUAS/ICI, calculation of 'proxy-Z-scores' was not necessary.

7.2 Assigned value

For EQUAS assessment, the concentration determined by the expert laboratories is used as assigned value. Therefore, at least three expert laboratories must provide assessable results. For each expert laboratory the individual means and standard deviations were calculated. The mean of all individual means (mean-of-means, mom) and its standard deviation (SD_{mom}) were calculated with the relative uncertainty given in Equation 1.

$$u = \frac{SD_{mom}}{\sqrt{n}} \quad \text{Equation 1}$$

u uncertainty of the mean of the mean concentrations from the expert labs
 SD_{mom} standard deviation of the mean of the mean concentrations
n number of expert laboratories (after exclusion of outliers if applicable)

The mean of all expert laboratories is considered suitable as assigned value for EQUAS studies, if $u \leq 0.7 \sigma_T$ (σ_T see section 7.3). When $u > 0.7 \sigma_T$, individual means are checked for outliers. Extreme outliers that are outside the range of $\pm 50 \%$ of the median are excluded from the data set. If exclusion of outliers doesn't yield into reliable results and u is still $> 0.7 \sigma_T$, uncertainty of the expert-derived mean is too high to be used as assigned value. In this case, no EQUAS assessment of participants' performance is possible for the applicable biomarker.

When no EQUAS assessment of the participants' performance is possible or the number of expert laboratories is not sufficient for EQUAS assessment, the consensus value is used as the assigned value and calculated as described in HBM4EU-SOP-QU-003. In brief, the consensus value and its uncertainty is calculated from the results submitted by the participants using robust statistic in order to minimize the influence of outliers. Extreme outliers that are outside the range of $\pm 50 \%$ of the median are excluded from the data set. The robust mean is taken as consensus value. The uncertainty of the consensus values is calculated according to Equation 2. For calculating Z-scores two criteria were evaluated:

- (1) If $u \leq 0.3 \sigma_T$, the uncertainty of the consensus value may be considered negligible, otherwise another performances score must be calculated taken into account the uncertainty.
- (2) If $u \leq 0.7 \sigma_T$, the uncertainty of the consensus value is within acceptable limits and Z-scores can be provided.

$$u = 1.25 \frac{\hat{\sigma}}{\sqrt{n}} \quad \text{Equation 2}$$

u uncertainty of the consensus value
 $\hat{\sigma}$ standard deviation of the participants' results
n number of results used for calculating the consensus value

7.3 Target standard deviation (σ_T)

For calculation of the Z-scores, a fit-for-purpose target standard deviation (FFP) of 25 % of the assigned value is used as target standard deviation (σ_T). This was the default standard indicated in HBM4EU-SOP-QA-003.

7.4 Z-scores

Z-scores are calculated according to SOP HBM4EU-SOP-QA-003 using Equation 3.

$$Z = \frac{x-C}{\sigma_T} \quad \text{Equation 3}$$

Z	Z-score for submitted analysis result
x	result submitted by the laboratory
C	expert assigned value and consensus value, respectively
σ_T	target standard deviation, here 0.25 C

In accordance with ISO 13528 and ISO 17043, Z-scores are classified as presented in Table 4.

Table 4: Classification of Z-Scores

$ Z \leq 2$	Satisfactory
$2 < Z < 3$	Questionable
$ Z \geq 3$	Unsatisfactory

8 Results and Discussion

8.1 Participants' Results

In total, nine laboratories from four countries agreed to participate in this study, of which eight submitted results. Appendix VII (**Table 8-Table 9**) gives an overview of the results submitted by the participants. None of the laboratories reported results labelled as 'not detected' or <LOQ. In addition, no participant detected a false positive or negative result.

8.2 Assigned values and (target) standard deviations

For the following anilines, the results from the expert laboratories were used for calculating the mean, which was used as the assigned value: **2,4-TDA** (low and high), **2,6-TDA** (low), **MOCA** (high), **TOL** (low and high), **MDA** (low and high).

For **MOCA** (low), the consensus value of the ICI was used as assigned value, since uncertainty of the expert laboratories was too high ($u > 0.7 \sigma_T$) to calculate an assigned value.

In case of **2,6-TDA** (high), uncertainty within the four expert laboratories was also too high ($u > 0.7 \sigma_T$) for EQUAS assessment. Exclusion of outliers did not lead to reliable results and thus the consensus value (ICI) was calculated. However, variation within the seven participants required the exclusion of outliers, leaving

only five laboratories for calculating the consensus value. Nevertheless, it was used as assigned value, since the total number of submitted results after exclusion of outliers ($N = 28$) was considered sufficient.

For **AN** (low and high), EQUAS requirements after exclusion of outliers did not meet the criterion ($u > 0.7 \sigma_T$) and a consensus value could not be calculated due to the lack of participants. Thus, no assigned value could be determined.

Assigned consensus value (EQUAS or ICI) and its uncertainty, standard deviation (SD), fit-for-purpose (FFP) relative target standard deviation (σ_T) and study RSD_R of each analyte are given in Appendix VIII.

8.3 Assessment of laboratory performance

An overview of the number of laboratories with satisfactory, questionable, and unsatisfactory Z-scores for the respective aromatic amines as well as the evaluation scheme applied can be found in Table 1, Chapter 1. More detailed results of the laboratory performance for the respective biomarker in the low and high concentration level is given in Table 2, Chapter 1.

Z-Scores could be calculated for the low and high concentration level of 2,4-TDA, 2,6-TDA, MDA, MOCA, and TOL. Graphical evaluation of the laboratories' performance is given in Appendix IX (Figure 1-Figure 5). In general, a satisfactory performance was achieved for TOL (100 % for TOL_{Low} and TOL_{High}) and MDA (100 % for MDA_{Low} and 88 % for MDA_{High}). High variations within the data set were observed for $MOCA_{High}$ with 63 % satisfactory results. For AN (low and high), no Z-Scores could be provided, since a calculation of an assigned value was not possible due to the high variability of the results derived from the expert laboratories and an insufficient number ($N < 7$) of participating laboratories for Z-score calculation.

8.4 Conclusion and recommendations

The HBM4EU EQUAS/ICI Round 1 was performed successfully, however, the total number of participating labs was quite low compared to other HBM4EU EQUAS/ICI programs. Eight of nine registered laboratories reported results. Table 5 gives an overview of the laboratories' performance and the LOQ for the corresponding aromatic amines in urine. Evaluation of laboratory performance was possible for five aromatic amines: 2,4-TDA, 2,6-TDA, MDA, MOCA, and TOL. Percentage of satisfactory Z-scores was between 63 % and 100 %. The submitted results of AN were not sufficient for an evaluation and therefore this analyte does not appear suitable for an EQUAS/ICI.

Table 5: Performance of the participating laboratories for aromatic amines in urine

Laboratory code	LOQ [ng/mL]	2,4-TDA _{Low}	2,4-TDA _{High}
AA_01	25	Satisfactory	Satisfactory
AA_03	1.0	Unsatisfactory	Unsatisfactory
AA_05	NA	NA	NA
AA_06	NA	NA	NA
AA_07	0.3	Satisfactory	Satisfactory
AA_10	1.0	Satisfactory	Satisfactory
AA_12	0.6	Satisfactory	Satisfactory
AA_16	0.1	Questionable	Questionable
AA_21	0.2	Satisfactory	Satisfactory

		2,6-TDA _{Low}	2,6-TDA _{High}
AA_01	25	Satisfactory	Satisfactory
AA_03	1.0	Unsatisfactory	Unsatisfactory
AA_05	NA	NA	NA
AA_06	NA	NA	NA
AA_07	0.3	Satisfactory	Satisfactory
AA_10	1.0	Satisfactory	Satisfactory
AA_12	0.6	Satisfactory	Satisfactory
AA_16	0.1	Unsatisfactory	Questionable
AA_21	0.2	Satisfactory	Satisfactory

		AN _{Low}	AN _{High}
AA_01	0.2	NE	NE
AA_03	0.5	NA	NA
AA_05	2.0	NE	NE
AA_06	NA	NA	NA
AA_07	NA	NA	NA
AA_10	NA	NA	NA
AA_12	NA	NA	NA
AA_16	0.1	NE	NE
AA_21	2.0	NE	NE

Table 5 continued

Laboratory code	LOQ [ng/mL]	MDA _{Low}	MDA _{High}
AA_01	1.0	Satisfactory	Satisfactory
AA_03	0.1	Satisfactory	Satisfactory
AA_05	NA	NA	NA
AA_06	1.0	Satisfactory	Satisfactory
AA_07	0.3	Satisfactory	Satisfactory
AA_10	0.5	Satisfactory	Satisfactory
AA_12	1.0	Satisfactory	Satisfactory
AA_16	0.0	Satisfactory	Questionable
AA_21	0.1	Satisfactory	Satisfactory

		MOCA _{Low}	MOCA _{High}
AA_01	10	Questionable	Satisfactory
AA_03	0.1	Satisfactory	Questionable
AA_05	NA	NA	NA
AA_06	5.5	Satisfactory	Satisfactory
AA_07	0.3	Satisfactory	Satisfactory
AA_10	1.3	Satisfactory	Questionable
AA_12	1.3	Satisfactory	Satisfactory
AA_16	0.1	Unsatisfactory	Unsatisfactory
AA_21	0.2	Satisfactory	Satisfactory

		TOL _{Low}	TOL _{High}
AA_01	0.01	Satisfactory	Satisfactory
AA_03	NA	NA	NA
AA_05	NA	NA	NA
AA_06	NA	NA	NA
AA_07	NA	NA	NA
AA_10	0.60	Satisfactory	Satisfactory
AA_12	NA	NA	NA
AA_16	0.10	Satisfactory	Satisfactory
AA_21	0.05	Satisfactory	Satisfactory

With $|Z| \leq 2$ – Satisfactory, $2 < |Z| < 3$ – Questionable, $|Z| \geq 3$ – Unsatisfactory.

9 References

HBM4EU-SOP-QA-001 'Organisation of Interlaboratory Comparison Investigations (ICI) and External Quality Assurance Schemes (EQUAS) of interlaboratory studies'

HBM4EU-SOP-QA-002 'Preparation of test materials for ICI / EQUAS'

HBM4EU-SOP-QA-003 'Evaluation of ICI / EQUAS results'

HBM4EU-SOP-QA-004 'Reporting of ICI / EQUAS studies'

ISO/IEC 17043:2010, Conformity assessment – General requirements for proficiency testing

ISO 13528, 2015, Statistical methods for use in proficiency testing by interlaboratory comparison.

Fearn, T. and M. Thompson (2001). *A new test for 'sufficient homogeneity'*. Analyst 126(8): 1414-1417.

Thompson, M., S. L. Ellison and R. Wood (2006). *The international harmonized protocol for the proficiency testing of analytical chemistry laboratories (IUPAC Technical Report)*. Pure and Applied Chemistry 78(1): 145-196.

Appendix

I Homogeneity data

Table 6: Homogeneity data of 2,4-TDA, 2,6-TDA, Aniline, MDA, MOCA, and o-Toluidine for low and high concentration range.

	2,4-TDA				2,6-TDA				Aniline			
	Low [ng/mL]		High [ng/mL]		Low [ng/mL]		High [ng/mL]		Low [ng/mL]		High [ng/mL]	
	Replicate 1	Replicate 2	Replicate 1	Replicate 2	Replicate 1	Replicate 2	Replicate 1	Replicate 2	Replicate 1	Replicate 2	Replicate 1	Replicate 2
	44.8	53.9	181.0	189.0	52.5	48.6	212.0	242.0	2.7	2.3	11.5	12.0
	49.2	64.3	185.0	189.0	55.4	52.7	193.0	204.0	3.0	2.1	11.3	12.5
	54.8	56.3	194.0	205.0	56.4	50.7	208.0	225.0	2.9	2.2	11.7	11.6
	49.0	54.5	174.0	206.0	54.5	56.4	214.0	216.0	3.0	2.9	11.6	11.6
	49.2	60.2	194.0	194.0	52.3	55.5	198.0	213.0	3.1	2.1	11.8	11.3
	52.8	50.8	197.0	196.0	47.9	57.1	224.0	207.0	2.8	3.1	11.9	11.4
	53.8	55.7	200.0	198.0	45.1	51.5	212.0	220.0	2.8	3.0	11.9	11.3
	47.8	52.7	190.0	214.0	57.3	52.8	196.0	211.0	3.1	3.0	11.7	11.4
	51.2	57.8	209.0	194.0	50.4	48.6	219.0	210.0	2.7	2.6	11.8	12.7
	52.6	51.3	195.0	207.0	51.2	48.7	195.0	238.0	3.0	2.2	12.1	12.1
Grand mean	53.1		195.6		52.3		212.9		2.7		11.8	
Cochran's test												
C	0.4		0.5		0.4		0.5		0.3		0.4	
C _{crit}	0.6		0.6		0.6		0.6		0.6		0.6	
Detection of outliers (C > C _{crit})	no outliers detected		no outliers detected		no outliers detected		no outliers detected		no outliers detected		no outliers detected	
target σ _{FFP}	13.3		48.9		13.1		53.2		0.7		2.9	
s _x	2.4		6.3		2.5		8.2		0.2		0.2	
s _w	5.2		10.4		3.3		14.2		0.4		0.4	
s _s	0.0		0.0		0.9		0.0		0.0		0.0	
C ₁ = 0.3 σ _{FFP}	4.0		14.7		3.9		16.0		0.2		0.9	
s _s < C ₁ ?	Yes, homogeneity adequate		Yes, homogeneity adequate		Yes, homogeneity adequate		Yes, homogeneity adequate		Yes, homogeneity adequate		Yes, homogeneity adequate	
s _w < 0.5 σ _{FFP} ?	Yes, method suitable		Yes, method suitable		Yes, method suitable		Yes, method suitable		No, method not suitable		Yes, method suitable	

Table 6 continued

	MDA				MOCA				o-Toluidine			
	Low [ng/mL]		High [ng/mL]		Low [ng/mL]		High [ng/mL]		Low [ng/mL]		High [ng/mL]	
	Replicate 1	Replicate 2	Replicate 1	Replicate 2	Replicate 1	Replicate 2	Replicate 1	Replicate 2	Replicate 1	Replicate 2	Replicate 1	Replicate 2
	4.7	5.2	97.0	94.8	15.4	10.8	148.0	136.0	0.32	0.32	1.51	1.47
	5.7	5.2	97.3	98.7	15.3	20.9	184.0	152.0	0.32	0.30	1.46	1.54
	4.9	5.1	109.0	105.0	19.2	13.3	105.0	147.0	0.31	0.31	1.54	1.46
	4.7	5.3	94.0	104.0	15.0	11.3	113.0	150.0	0.30	0.33	1.44	1.50
	5.1	5.8	100.0	98.7	11.5	9.8	132.0	135.0	0.33	0.31	1.46	1.39
	4.9	5.4	103.0	96.2	15.5	8.4	103.0	112.0	0.31	0.32	1.50	1.45
	5.1	5.7	85.1	97.8	12.3	12.1	128.0	124.0	0.31	0.32	1.51	1.44
	5.4	5.0	103.0	97.5	14.3	16.4	151.0	155.0	0.32	0.31	1.49	1.47
	4.9	5.1	105.0	103.0	18.4	12.8	125.0	130.0	0.32	0.31	1.45	1.57
	5.1	5.0	91.4	97.6	8.7	12.7	133.0	200.0	0.32	0.30	1.51	1.48
grand mean	5.2		98.9		13.7		138.2		0.32		1.48	
Cochran's test												
C	0.2		0.4		0.2		0.5		0.20		0.33	
C _{crit}	0.6		0.6		0.6		0.6		0.60		0.60	
Detection of outliers (C > C _{crit})	no outliers detected		no outliers detected		no outliers detected		no outliers detected		no outliers detected		no outliers detected	
target σ _{FFP}	1.3		24.7		3.4		34.5		0.08		0.37	
s _x	0.2		4.5		2.5		19.3		0.00		0.02	
s _w	0.4		4.5		3.2		21.1		0.01		0.05	
s _s	0.0		3.1		1.1		12.2		0.00		0.00	
C ₁ = 0.3 σ _{FFP}	0.4		7.4		1.0		10.4		0.02		0.11	
s _s < C ₁ ?	Yes, homogeneity adequate		Yes, homogeneity adequate		No, homogeneity inadequate		No, homogeneity inadequate		Yes, homogeneity adequate		Yes, homogeneity adequate	
s _w < 0.5 σ _{FFP} ?	Yes, method suitable		Yes, method suitable		No, method not suitable		No, method not suitable		Yes, method suitable		Yes, method suitable	
With additional allowance for sampling errors and repeatability												
s _{sam} ²					1.1		148.1					
C ₂ = F ₁ σ _{all} ² + F ₂ s _{an} ²					12.4		653.1					
s _{sam} ² < C ₂ ?					Yes, homogeneity adequate		Yes, homogeneity adequate					

Replicate	2,4-TDA				2,6-TDA				Aniline			
	Low [ng/mL]		High [ng/mL]		Low [ng/mL]		High [ng/mL]		Low [ng/mL]		High [ng/mL]	
	t = 0 d	t = 46 d	t = 0 d	t = 46 d	t = 0 d	t = 46 d	t = 0 d	t = 46 d	t = 0 d	t = 44 d	t = 0 d	t = 44 d
1	44.8	52.1	181.0	193.0	52.5	50.7	212.0	206.0	2.7	3.3	11.5	11.1
2	49.2	54.1	185.0	187.0	55.4	53.7	193.0	204.0	3.0	3.2	11.3	11.4
3	54.8	51.3	194.0	174.0	56.4	47	208.0	199.0	2.9	3.0	11.7	11.6
Average	49.6	52.5	186.7	184.7	54.8	50.5	204.3	203.0	2.8	3.2	11.5	11.3
SD	5.0	1.4	6.7	9.7	2.0	3.4	10.0	3.6	0.2	0.2	0.2	0.2
Difference (x_0-x_E)	-2.9		2.0		4.3		1.3		-0.3		0.2	
$c = 0.3 \sigma_{FFP}$	3.7		14.0		4.1		15.3		0.2		0.9	
$(x_0-x_E) < c?$	No		No		Yes		No		Yes		No	
Consequential instability												
t	1.0		0.3		1.9		0.2		2.3		0.9	
t _{Crit}	2.8		2.8		2.8		2.8		2.8		2.8	
Significant difference	No		No		No		No		No		No	

Table 7 continued

Replicate	MDA				MOCA				o-Toluidine			
	Low [ng/mL]		High [ng/mL]		Low [ng/mL]		High [ng/mL]		Low [ng/mL]		High [ng/mL]	
	t = 0 d	t = 46 d	t = 0 d	t = 46 d	t = 46 d	t = 46 d	t = 0 d	t = 46 d	t = 0 d	t = 44 d	t = 0 d	t = 44 d
1	4.66	6.36	97.0	96.8	15.4	12.1	148.0	166.0	0.3	0.3	1.5	1.4
2	5.73	5.48	97.3	101.0	15.3	17.2	184.0	134.0	0.3	0.3	1.5	1.4
3	4.86	5.7	109.0	87.3	19.2	12.3	105.0	133.0	0.3	0.3	1.5	1.4
Average	5.1	5.8	101.1	95.0	16.6	13.9	145.7	144.3	0.3	0.3	1.5	1.4
SD	0.6	0.5	6.8	7.0	2.2	2.9	39.6	18.8	0.0	0.0	0.0	0.0
Difference (x0-xE)	-0.8		6.1		2.8		1.3		0.0		0.1	
c = 0.3 σFFP	0.4		7.6		1.2		10.9		0.0		0.1	
(x0-xE) < c?	Yes		No		Yes		No		No		Yes	
Consequential instability	Yes		No		Yes		No		No		Yes	
t	1.8		1.1		1.3		0.1		4.6		5.1	
tcrit	2.8		2.8		2.8		2.8		2.8		2.8	
Significant difference	No		No		No		No		Yes		Yes	

III Method information form for the participation in EQUAS/ICI (Round 1)

I.I Method information for measuring homogeneity and stability of AN and TOL

Method information form for participation in ICI/EQUAS Aromatic Amines in urine/Round 1

Laboratory code		
ISO17025 accredited	Yes	
SAMPLE PREPARATION		
amount sample extracted	5 mL	mL
Extraction		
- pH adjustment	6-6.4	
- LLE;	2x 5 mL Hexane (15 min)	
- SPE; material	--	
Cleanup		
- LLE; solvent(s)	--	
- SPE; material	--	
Derivatisation		
- reagent	PFPA	
INSTRUMENTAL ANALYSIS		
HPLC		
- injection volume	--	
- column stationary phase	--	
- column L (mm) x ID (mm); dp (µm)	--	
- temperature	--	
- mobile phase A	--	
- mobile phase B	--	
- flow rate	--	
GC		
- injector	Split (1:15)	
- injection volume	2 µL	µL
- column stationary phase	RXI-5ms	
- column L (m) x ID (mm) df (µm)	30 m, 0.25 mm ID, 0.25 µm df	
- carrier	He	
- flow rate / inlet pressure	1.1 mL/min, constant flow	

Detection		
MS	single quad	
other	--	
Quantification		
Use of internal standard (IS)	Yes	
- isotopic label	Yes	
- other	--	
- moment of addition	before deconjugation	
- response normalised to IS	yes	
Calibration	matrix-matched (addition to blank matrix before extraction)	
	multi level	
Correction for recovery	No	
Identification criteria used		
- retention time tolerance	identification via retention time of isotopically labelled std.	
- number of ions/transitions	--	
- ion ratio tolerance	--	

Further remarks/observations:

Date:

Signature:

I.II Method information for measuring homogeneity and stability of 2,4-TDA, 2,6-TDA, MDA, and MOCA

Method information form for participation in ICI/EQUAS Aromatic Amines in urine/Round 1

Laboratory code	QR/122	
ISO17025 accredited	No	
SAMPLE PREPARATION		
amount sample extracted	0.25	g or <u>mL</u>
Extraction		
- pH adjustment	pH 1	
- LLE;	solvent(s) / time / shaking	
- SPE; material	Material Strata XC	
Cleanup		
- LLE; solvent(s)	--	
- SPE; material	--	
Derivatisation		
- reagent	--	
INSTRUMENTAL ANALYSIS		
HPLC		
- injection volume	5	μ L
- column stationary phase	Gemini C18	
- column L (mm) x ID (mm); dp (μ m)	150 x 3 mm; 3 μ m	
- temperature	35°C	
- mobile phase A	NH ₄ OAc 5 mM; pH 9.2	
- mobile phase B	ACN + 5 mM NH ₄ OAc	
- flow rate	1	mL/min
GC		
- injector	splitless/PTV/....	
- injection volume	--	
- column stationary phase	--	
- column L (m) x ID (mm) df (μ m)	--	
- carrier	--	
- flow rate / inlet pressure	--	

Detection		
MS	Triple Quad	
other	--	
Quantification		
Use of internal standard (IS)	Yes	
- isotopic label	Yes	
- other	specify	
- moment of addition	before acid hydrolysis	
- response normalised to IS	Yes	
Calibration	matrix-matched (addition to blank matrix before extraction)	
	<u>multi level</u>	
Correction for recovery	No	
Identification criteria used		
- retention time tolerance	Coherence with IS	
- number of ions/transitions	2	
- ion ratio tolerance	% relative/absolute deviation from reference standard	

Further remarks/observations:

Date:

Signature:

IV Copy of invitation letter

HBM4EU: Announcement / invitation to participate in ICI / EQUAS study

Anilines/Round 1

Title of ICI/EQUAS: Anilines in urine

Dear Colleague,

within the frame of HBM4EU the

ABF GmbH
Simmelweisstr. 5
82152 Planegg
Germany

announces the 1st round of ICI/EQUAS for the determination of anilines (aromatic amines) in urine.

The aim of ICI/EQUAS exercises is to provide laboratories with an assessment of their analytical performance and reliability of their data in comparison with other laboratories and/or expert laboratories. This will aid in the quality improvement of analysis in human biomonitoring at each of the laboratories.

Participation is mandatory for laboratories analysing samples in the frame of HBM4EU.

Participants meet the quality criteria of the HBM4EU call if they pass 3 of ICI/EQUAS rounds successfully.

Test samples

The matrix will be urine. Accordingly, the participants will receive:

- 2 different materials (levels) of urine (3 samples per level of 8mL each) for determination of anilines/aromatic amines in urine

Target biomarkers

For the biomarkers potentially present in the test samples, please see registration form for Anilines/Round 1. We would be pleased if your laboratory could analyse as many metabolites as possible.

LOQs should allow the analysis of aromatic amines in samples of the general population.

Calendar:

Receipt of test samples	29-05-2019 – 31-05-2019
Deadline for submission of results	05-07-2019

Fee

For partners and linked-third parties of HBM4EU, participation is free of charge. Please note that the participant is responsible for custom clearance and associated costs if applicable.

Confidentiality:

All laboratory-specific information will be treated confidentially, and will never be disclosed to third parties (government, accreditation bodies) except the HBM4EU QAU, without permission of the laboratory.

Contact information organiser:

Coordinators:

- Dr. Nikola Pluym
- Dr. Max Scherer

ABF GmbH
Simmelweisstr. 5
82152 Planegg
Germany

E-mail: nikola.pluym@abf-lab.com
max.scherer@abf-lab.com

V Copy of registration form

HBM4EU: Registration form for participation in ICI / EQUAS study Aniline /Round 1

Title of ICI/EQUAS: Anilines in urine

Please choose the metabolites you want to participate with.

We would appreciate your registration for as many biomarkers as possible.

Parameter	Participation	
	yes	no
4,4'-MDA		
MOCA		
aniline		
p-aminophenol		
N-acetyl-4-aminophenol		
p-PDA		
o-toluidine		
2,4-TDA		
2,6-TDA		

Participating laboratory:

name of the institution

address of the laboratory

name of 1st contact person, telephone number and e-mail address

name of 2nd contact person, telephone number and e-mail address

Address for delivery of the test samples

name of the institution

address of the laboratory

The above laboratory will participate in the ICI/EQUAS study **Anilines/Round 1**.

I agree with the conditions mentioned in the invitation letter, and that the laboratory will analyse the ICI/EQUAS samples using the same procedure as will be used for analysis of samples in the frame of HBM4EU, and submit results before the indicated deadline.

Name:

Signature:

Date:

After signing this form, please scan and send the pdf to:

nikola.pluym@abf-lab.com
max.scherer@abf-lab.com

The registration deadline is 21-12-2018.

Contact information organiser:

- Nikola Pluym
- Max Scherer
- Gerhard Scherer

ABF GmbH
Sammelweisstr. 5
82152 Planegg
Germany

E-mail: nikola.pluym@abf-lab.com

VI Copy of instruction letter

Instructions for ICI / EQUAS study Anilines (aromatic amines) in urine Round 01/2019 sent to the participants by e-mail

Dear Colleague,

we have dispatched the samples yesterday frozen on dry-ice. You shall receive the samples by tomorrow Wednesday 29-May.

Each lab will receive 3 samples per level (2 levels). Expert labs will receive 6 samples per level (expert labs please use the corresponding form). The levels are labeled as A and B with a unique aliquot number. Please send us a short e-mail notification after sample receipt on the conditions of the samples (frozen? / leakage?), number of samples per aliquot (A/B) and date of receipt.

Regarding sample analyses please consider the following.

Due to the poor stability of the free analytes we prepared the urine material by spiking with the conjugates wherever these were available. We ask you therefore to perform your complete sample preparation procedure including the hydrolysis step.

We are aware that some of the laboratories will need to run two methods for quantification of the parameters. Since we were limited in the volume of urine (8 mL per aliquot) we prepared the samples in a manner that shall allow analysis after dilution with respect to the reported LOQs of the most commonly used assays in the literature (we assumed the following LOQs from the literature: 1 µg/L for diamines 2,4-TDA; 2,6-TDA; MDA; MOCA; and LOQs of 50 ng/L for o-toluidine and 500 ng/L for aniline. For these LOQs a dilution by at least 3 to 5-fold shall be feasible).

I hope this information is of help in planning the analyses. I am sure that you understand that I cannot give you much more information about the expected levels.

For reporting, please use the result sheet (Excel) and follow the instructions of the result submission form. We also ask you to fill out the method information form.

You will receive your laboratory code in a separate e-mail in the following days.

Please be reminded that the deadline for reporting of the data is the 05-July-2019.

Please do not hesitate to contact us if you have any questions in the meantime.

Dr. Nikola Pluym
Phone +49-89-535395
E-mail: nikola.pluym@abf-lab.com

VII Results of the control material analyzed by the participants

Table 8: Results for 2,4-TDA, 2,6-TDA, and AN of all participating laboratories

Lab Code	Sample No.	2,4-TDA [ng/mL]			2,6-TDA [ng/mL]			AN [ng/mL]		
		Low	High	LOQ	Low	High	LOQ	Low	High	LOQ
AA_01	1	44.8	181.0		52.5	212.0		2.7	11.5	
	2	49.2	185.0		55.4	193.0		3.0	11.3	
	3	54.8	194.0		56.4	208.0		2.9	11.7	
	4	49.0	174.0		54.5	214.0		3.0	11.6	
	5	49.2	194.0	25	52.3	198.0	25	3.1	11.8	0.2
	6	52.8	197.0		47.9	224.0		2.8	11.9	
	7	53.8	200.0		45.1	212.0		2.8	11.9	
	8	47.8	190.0		57.3	196.0		3.1	11.7	
	9	51.2	209.0		50.4	219.0		2.7	11.8	
	10	52.6	195.0		51.2	195.0		3.0	12.1	
AA_12	1	48.0	153.7		44.5	124.6		NA	NA	
	2	48.4	153.0		44.2	131.2		NA	NA	
	3	46.7	149.4	0.6	42.1	127.6	0.6	NA	NA	NA
	4	48.0	149.5		43.9	120.3		NA	NA	
	5	48.2	147.1		44.3	120.4		NA	NA	
	6	49.0	157.3		43.9	133.4		NA	NA	
AA_16	1	18.0	74.5		14.1	64.3		0.2	0.6	
	2	15.2	68.1		12.0	58.0		0.2	0.5	
	3	12.5	73.0	0.1	10.0	58.4	0.1	0.2	0.6	0.1
	4	16.7	68.0		11.7	55.8		0.2	0.5	
	5	20.2	63.6		12.5	53.6		0.2	0.6	
	6	19.7	62.7		13.8	51.4		0.2	0.6	
AA_21	1	47.2	190.7		59.3	265.3		51.3	46.6	
	2	47.6	182.0		57.8	252.1		50.4	54.0	
	3	48.3	176.7	0.2	58.9	247.8	0.2	44.3	47.1	2.0
	4	48.4	180.8		59.5	254.0		47.2	52.4	
	5	48.0	182.7		58.7	253.7		49.1	50.9	
	6	48.2	180.9		58.9	252.4		50.7	45.7	
AA_03	1	2.5	14.6		1.0	9.0		NA	NA	
	2	4.3	14.6	1.0	1.1	7.0	1.0	NA	NA	0.5
	3	4.0	17.2		1.2	7.9		NA	NA	
AA_05	1	NA	NA		NA	NA		24.6	34.5	
	2	NA	NA	NA	NA	NA	NA	25.1	34.3	2.0
	3	NA	NA		NA	NA		23.3	32.9	
AA_06	1	NA	NA		NA	NA		NA	NA	
	2	NA	NA	NA	NA	NA	NA	NA	NA	NA
	3	NA	NA		NA	NA		NA	NA	
AA_07	1	55.2	219.4		51.2	195.7		NA	NA	
	2	55.2	218.8	0.3	49.8	196.7	0.3	NA	NA	NA
	3	55.1	219.6		49.8	205.2		NA	NA	
AA_10	1	42.9	145.0		44.4	191.0		NA	NA	
	2	45.8	146.0	1.0	48.5	166.0	1.0	NA	NA	NA
	3	46.8	149.0		48.7	184.0		NA	NA	

NA – not analyzed

Table 9: Results for MDA, MOCA, and TOL of all participating laboratories

Lab Code	Sample No.	MDA [ng/mL]			MOCA [ng/mL]			TOL [ng/mL]		
		Low	High	LOQ	Low	High	LOQ	Low	High	LOQ
AA_01	1	4.7	97.0		15.4	148.0		0.32	1.51	
	2	5.7	97.3		15.3	184.0		0.32	1.46	
	3	4.9	109.0		19.2	105.0		0.31	1.54	
	4	4.7	94.0		15.0	113.0		0.30	1.44	
	5	5.1	100.0	1.0	11.5	132.0	10	0.33	1.46	0.01
	6	4.9	103.0		15.5	103.0		0.31	1.50	
	7	5.1	85.1		12.3	128.0		0.31	1.51	
	8	5.4	103.0		14.3	151.0		0.32	1.49	
	9	4.9	105.0		18.4	125.0		0.32	1.45	
	10	5.1	91.4		8.7	133.0		0.32	1.51	
AA_12	1	5.7	94.8		8.5	81.0		NA	NA	
	2	5.5	90.7		8.8	78.0		NA	NA	
	3	5.1	85.4	1.0	7.6	73.3	1.3	NA	NA	NA
	4	5.6	87.8		7.3	78.2		NA	NA	
	5	5.4	90.0		8.7	78.4		NA	NA	
	6	5.6	95.2		8.3	79.6		NA	NA	
AA_16	1	4.2	31.5		0.1	0.1		0.30	1.27	
	2	5.3	34.4		0.1	0.1		0.27	1.28	
	3	4.3	35.6	0.0	0.1	0.1	0.1	0.28	1.27	0.10
	4	3.1	30.2		0.1	0.1		0.28	1.08	
	5	4.1	29.5		0.1	0.1		0.28	1.26	
	6	4.6	31.0		0.1	0.1		0.29	1.31	
AA_21	1	4.5	95.5		11.3	95.3		0.36	1.41	
	2	4.4	94.1		10.2	109.4		0.35	1.41	
	3	4.6	93.5	0.1	10.0	109.7	0.2	0.35	1.44	0.05
	4	4.2	93.9		8.5	108.8		0.35	1.44	
	5	4.5	95.2		9.5	115.2		0.35	1.45	
	6	4.7	94.5		9.8	113.8		0.35	1.42	
AA_03	1	2.0	68.5		5.1	50.6		NA	NA	
	2	3.4	43.6	0.1	5.3	50.3	0.1	NA	NA	NA
	3	2.1	64.9		5.0	49.5		NA	NA	
AA_05	1	NA	NA		NA	NA		NA	NA	
	2	NA	NA	NA	NA	NA	NA	NA	NA	NA
	3	NA	NA		NA	NA		NA	NA	
AA_06	1	5.0	78.6		5.3	65.9		NA	NA	
	2	4.5	82.4	1.0	5.2	66.3	5.5	NA	NA	NA
	3	4.5	78.7		6.6	63.2		NA	NA	
AA_07	1	4.7	88.7		10.3	83.9		NA	NA	
	2	4.8	88.7	0.3	11.0	90.3	0.3	NA	NA	NA
	3	4.7	88.3		10.7	88.4		NA	NA	
AA_10	1	3.4	73.1		7.6	47.7		0.25	0.99	
	2	3.7	69.9	0.5	7.8	50.6	1.3	0.25	0.94	0.60
	3	3.8	72.9		8.2	52.7		0.24	1.03	

NA – not analyzed

VIII Assigned values and participants' performance for Anilines in urine

Table 10: Assigned value and participants' performance (mean) for 2,4-Diaminotoluene (2,4-TDA) in urine

	2,4-TDA _{Low}	2,4-TDA _{High}		
Assigned value from 3 expert labs (low and high)	49.1 ng/mL	178.3 ng/mL		
Standard deviation (SD) expert SD (low and high)	2.4 ng/mL	18.6 ng/mL		
uncertainty of assigned value (u)	1.3 %	2.8 %		
target standard deviation (σ _T , 25 %)	12.3 ng/mL	44.6 ng/mL		
Relative target standard deviation (%)	25 %	25 %		
0.7 · σ _T	17.5 %	17.5 %		
study RSD _R	34.0 %	34.2 %		
Laboratory code	2,4-TDA [ng/mL]	Z-Score	2,4-TDA [ng/mL]	Z-Score
AA_01	50.5	0.1	191.9	0.3
AA_03	3.6	-3.7	15.5	-3.7
AA_05	NA	NE	NA	NE
AA_06	NA	NE	NA	NE
AA_07	55.2	0.5	219.3	0.9
AA_10	45.2	-0.3	146.7	-0.7
AA_12	48.0	-0.1	151.7	-0.6
AA_16	17.0	-2.6	68.3	-2.5
AA_21	48.0	-0.1	182.3	0.1

NA – not analyzed, NE – not evaluable

Table 11: Assigned value and participants' performance for 2,6-Diaminotoluene (2,6-TDA) in urine

	2,6-TDA _{Low}	2,6-TDA _{High}
Assigned value assigned value from 3 expert labs (low)/ consensus value (high)	51.8 ng/mL	196.2 ng/mL **
Standard deviation (SD) expert SD (low)/ robust SD (high)	6.3 ng/mL	44.4 ng/mL
uncertainty of assigned value (u)	3.2 %	5.4 %
target standard deviation (σ _T , 25 %)	12.9 ng/mL	49.1 ng/mL
Relative target standard deviation (%)	25 %	25 %
0.7 · σ _T	17.5 %	17.5 %
study RSD _R	37.0 %	40.1 %

Laboratory code	2,6-TDA [ng/mL]	Z-Score	2,6-TDA [ng/mL]	Z-Score
AA_01	52.3	0.0	207.1	0.2
AA_03	1.1	-3.9	8.0	-3.8
AA_05	NA	NE	NA	NE
AA_06	NA	NE	NA	NE
AA_07	50.2	-0.1	199.2	0.1
AA_10	47.2	-0.4	180.3	-0.3
AA_12	43.8	-0.6	126.2	-1.4
AA_16	12.3	-3.0	56.9	-2.8
AA_21	58.9	0.5	254.2	1.2

NA – not analyzed, NE – not evaluable

** Consensus value calculated from five laboratories

Table 12: Assigned value and participants' performance for Aniline (AN) in urine

AN _{Low}			AN _{High}	
Assigned value			NE	
Laboratory code	Aniline [ng/mL]	Z-Score	Aniline [ng/mL]	Z-Score
AA_01	2.9	NE	11.7	NE
AA_03	NA	NE	NA	NE
AA_05	24.3	NE	33.9	NE
AA_06	NA	NE	NA	NE
AA_07	NA	NE	NA	NE
AA_10	NA	NE	NA	NE
AA_12	NA	NE	NA	NE
AA_16	0.2	NE	0.6	NE
AA_21	48.8	NE	49.5	NE

NA – not analyzed, NE – not evaluable

Table 13: Assigned value and participants' performance for 4,4'-Methylenedianiline (MDA) in urine

	MDA _{Low}	MDA _{High}
Assigned value assigned value from 4 (low) and 3 (high), respectively, expert labs	4.9 ng/mL	95.2 ng/mL
Standard deviation (SD) expert SD (low and high)	0.6 ng/mL	6.0 ng/mL
uncertainty of assigned value (u)	2.9 %	1.7 %
target standard deviation (σ_T , 25 %)	1.2 ng/mL	23.8 ng/mL
Relative target standard deviation (%)	25 %	25 %
$0.7 \cdot \sigma_T$	17.5 %	17.5 %
study RSD _R	18.0 %	24.8 %

Laboratory code	MDA [ng/mL]	Z-Score	MDA [ng/mL]	Z-Score
AA_01	5.0	0.2	98.5	0.1
AA_03	2.5	-1.9	59.0	-1.5
AA_05	NA	NE	NA	NE
AA_06	4.6	-0.2	79.9	-0.6
AA_07	4.7	-0.1	88.6	-0.3
AA_10	3.6	-1.0	72.0	-1.0
AA_12	5.5	0.5	90.6	-0.2
AA_16	4.3	-0.5	32.0	-2.7
AA_21	4.5	-0.3	94.5	0.0

NA – not analyzed, NE – not evaluable

Table 14: Assigned value and participants' performance for 4,4'-Methylenebis(2-chloroaniline) (MOCA) in urine

	MOCA _{Low}	MOCA _{High}
Assigned value consensus value (low)/ assigned value from 3 expert labs (high)	8.5 ng/mL	107.6 ng/mL
Standard deviation (SD) robust SD (low)/ expert SD (high)	2.1 ng/mL	23.5 ng/mL
uncertainty of assigned value (u)	6.0 %	6.0 %
target standard deviation (σ_T , 25 %)	2.2 ng/mL	26.9 ng/mL
Relative target standard deviation (%)	25 %	25 %
$0.7 \cdot \sigma_T$	17.5 %	17.5 %
study RSD _R	59.0 %	42.7 %

Laboratory code	MOCA [ng/mL]	Z-Score	MOCA [ng/mL]	Z-Score
AA_01	14.6	2.9	132.2	0.9
AA_03	5.1	-1.6	50.1	-2.1
AA_05	NA	NE	NA	NE
AA_06	5.7	-1.3	65.1	-1.6
AA_07	10.7	1.0	87.5	-0.7
AA_10	7.8	-0.3	50.3	-2.1
AA_12	8.2	-0.1	78.1	-1.1
AA_16	0.1	-4.0	0.1	-4.0
AA_21	9.9	0.7	108.7	0.0

NA – not analyzed, NE – not evaluable

Table 15: Assigned value and participants' performance for o-Toluidine (TOL) in urine

	TOL _{Low}	TOL _{High}
Assigned value from 3 expert labs (low and high)	0.32 ng/mL	1.40 ng/mL
Standard deviation (SD) expert SD (low and high)	0.03 ng/mL	0.11 ng/mL
uncertainty of assigned value (u)	2.2 %	2.1 %
target standard deviation (σ _T , 25 %)	0.08 ng/mL	0.35 ng/mL
Relative target standard deviation (%)	25 %	25 %
0.7 · σ _T	17.5 %	17.5 %
study RSD _R	11.0 %	12.5 %

Laboratory code	o-Toluidine [ng/mL]	Z-Score	o-Toluidine [ng/mL]	Z-Score
AA_01	0.32	0.0	1.49	0.2
AA_03	NA	NE	NA	NE
AA_05	NA	NE	NA	NE
AA_06	NA	NE	NA	NE
AA_07	NA	NE	NA	NE
AA_10	0.24	-0.9	0.99	-1.1
AA_12	NA	NE	NA	NE
AA_16	0.28	-0.4	1.25	-0.4
AA_21	0.35	0.4	1.43	0.1

NA – not analyzed, NE – not evaluable

IX Graphical evaluation of Z-Scores




Figure 1: Z-Scores of participating laboratories for the (A) low and (B) high concentration level of 2,4-TDA.




Figure 2: Z-Scores of participating laboratories for the (A) low and (B) high concentration level of 2,6-TDA.




Figure 3: Z-Scores of participating laboratories for the (A) low and (B) high concentration level of MDA.




Figure 4: Z-Scores of participating laboratories for the (A) low and (B) high concentration level of MOCA.




Figure 5: Z-Scores of participating laboratories for the (A) low and (B) high concentration level of TOL.